


BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071825\
Data File : VX047037.D
Acq On : 18 Jul 2025 07:32
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Title : SW846 8260
Last Update : Thu Jul 03 05:51:11 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.7	19659	PASS
75	95	30	60	52.5	52288	PASS
95	95	100	100	100.0	99603	PASS
96	95	5	9	6.7	6656	PASS
173	174	0.00	2	1.0	755	PASS
174	95	50	100	75.6	75264	PASS
175	174	5	9	7.3	5520	PASS
176	174	95	101	95.2	71645	PASS
177	176	5	9	6.7	4765	PASS