

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080823\
 Data File : VX036853.D
 Acq On : 08 Aug 2023 10:48
 Operator : JC/MD
 Sample : VX0808MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0808MBL01

Quant Time: Aug 09 04:10:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	139615	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	254490	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	239067	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	109036	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95519	46.238	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.480%
35) Dibromofluoromethane	5.385	113	80817	46.202	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.400%
50) Toluene-d8	8.647	98	306503	49.411	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.820%
62) 4-Bromofluorobenzene	11.079	95	118973	50.836	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080823\
Data File : VX036853.D
Acq On : 08 Aug 2023 10:48
Operator : JC/MD
Sample : VX0808MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0808MBL01

Quant Time: Aug 09 04:10:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

