

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080823\
 Data File : VX036860.D
 Acq On : 08 Aug 2023 13:39
 Operator : JC/MD
 Sample : 03915-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Aug 09 04:13:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	120085	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	219848	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	200919	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84865	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84817	47.735	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.460%
35) Dibromofluoromethane	5.391	113	70959	46.958	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.920%
50) Toluene-d8	8.647	98	262597	49.003	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.000%
62) 4-Bromofluorobenzene	11.079	95	96315	47.639	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080823\
Data File : VX036860.D
Acq On : 08 Aug 2023 13:39
Operator : JC/MD
Sample : 03915-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Aug 09 04:13:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

