

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080823\
 Data File : VX036863.D
 Acq On : 08 Aug 2023 14:51
 Operator : JC/MD
 Sample : 03903-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-03W

Quant Time: Aug 09 04:14:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	123070	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	226161	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209230	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94995	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86310	47.397	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.800%
35) Dibromofluoromethane	5.391	113	72597	46.701	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.400%
50) Toluene-d8	8.653	98	269349	48.860	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.720%
62) 4-Bromofluorobenzene	11.079	95	103561	49.793	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.580%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	8679	11.486	ug/l #	87
17) Carbon Disulfide	2.513	76	2515	0.661	ug/l	98
31) Cyclohexane	5.476	56	571	0.226	ug/l #	89
52) Toluene	8.726	92	1382	0.327	ug/l	91
67) Ethyl Benzene	10.195	91	93384	11.084	ug/l	98
68) m/p-Xylenes	10.305	106	3436	1.093	ug/l	98
69) o-Xylene	10.640	106	11374	3.625	ug/l	100
73) Isopropylbenzene	10.963	105	5358	0.682	ug/l	98
78) n-propylbenzene	11.305	91	6838	0.752	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	3612	0.549	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	22315	3.393	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080823\
Data File : VX036863.D
Acq On : 08 Aug 2023 14:51
Operator : JC/MD
Sample : 03903-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-03W

Quant Time: Aug 09 04:14:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

