

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080823\
 Data File : VX036875.D
 Acq On : 08 Aug 2023 19:38
 Operator : JC/MD
 Sample : 03903-07ME 10X
 Misc : 7.02g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-07ME

Quant Time: Aug 09 04:20:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	129269	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	237691	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	223121	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110010	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	90789	47.466	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.940%		
35) Dibromofluoromethane	5.385	113	75238	46.052	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.100%		
50) Toluene-d8	8.647	98	286321	49.420	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.840%		
62) 4-Bromofluorobenzene	11.079	95	114601	52.429	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	104.860%		
Target Compounds					Qvalue	
31) Cyclohexane	5.556	56	5900m	2.228	ug/l	
39) Methylcyclohexane	7.379	83	9494	3.135	ug/l	92
52) Toluene	8.720	92	17653	3.977	ug/l	98
67) Ethyl Benzene	10.195	91	608188	67.692	ug/l	99
68) m/p-Xylenes	10.299	106	957891	285.744	ug/l	97
69) o-Xylene	10.640	106	350118	104.633	ug/l	99
73) Isopropylbenzene	10.963	105	159725	17.544	ug/l	99
78) n-propylbenzene	11.305	91	624871	59.346	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	727324	95.494	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	2141690	281.171	ug/l	94
85) sec-Butylbenzene	11.890	105	75835	8.127	ug/l	96
86) p-Isopropyltoluene	12.006	119	31274	4.073	ug/l	99
89) n-Butylbenzene	12.329	91	184086	26.783	ug/l #	33

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080823\
Data File : VX036875.D
Acq On : 08 Aug 2023 19:38
Operator : JC/MD
Sample : 03903-07ME 10X
Misc : 7.02g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-07ME

Quant Time: Aug 09 04:20:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

