

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080823\
 Data File : VX036876.D
 Acq On : 08 Aug 2023 20:02
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 09 04:21:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	137495	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	251788	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	237116	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108508	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	97049	47.703	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	95.400%		
35) Dibromofluoromethane	5.385	113	79766	46.090	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.180%		
50) Toluene-d8	8.646	98	300358	48.940	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.880%		
62) 4-Bromofluorobenzene	11.079	95	117373	50.690	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	101.380%		
Target Compounds						
68) m/p-Xylenes	10.305	106	5095	1.430	ug/l	91
84) 1,2,4-Trimethylbenzene	11.749	105	11779	1.568	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080823\
Data File : VX036876.D
Acq On : 08 Aug 2023 20:02
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 09 04:21:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

