

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082224\
 Data File : VX043084.D
 Acq On : 22 Aug 2024 12:37
 Operator : JC/MD
 Sample : VSTDIC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Aug 23 04:18:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082224W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 04:08:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	369350	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	667190	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	551005	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	231473	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	7641	1.259	ug/l	0.01
Spiked Amount	50.000	Range	74 - 125	Recovery	=	2.520%#
35) Dibromofluoromethane	5.385	113	5829	1.266	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	2.540%#
50) Toluene-d8	8.647	98	18698	1.126	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	2.260%#
62) 4-Bromofluorobenzene	11.085	95	5798	1.003	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	2.000%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	2910	0.881	ug/l	92
3) Chloromethane	1.294	50	5657	1.142	ug/l	98
4) Vinyl Chloride	1.374	62	5644	1.076	ug/l #	87
5) Bromomethane	1.599	94	3168	1.013	ug/l	90
6) Chloroethane	1.666	64	4072	1.653	ug/l	97
7) Trichlorofluoromethane	1.867	101	9583	1.081	ug/l	91
8) Diethyl Ether	2.136	74	4064	1.060	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.313	101	4363	1.010	ug/l #	93
10) Methyl Iodide	2.441	142	4632	1.012	ug/l	89
11) Tert butyl alcohol	2.995	59	6225	4.950	ug/l #	91
12) 1,1-Dichloroethene	2.306	96	4548	1.048	ug/l	96
13) Acrolein	2.239	56	4819	Below Cal		98
14) Allyl chloride	2.660	41	8239	1.070	ug/l	96
15) Acrylonitrile	3.075	53	14507	4.914	ug/l	100
16) Acetone	2.392	43	14977	5.621	ug/l	98
17) Carbon Disulfide	2.502	76	11923	1.066	ug/l #	91
18) Methyl Acetate	2.709	43	9088	1.075	ug/l	97
19) Methyl tert-butyl Ether	3.117	73	15481	0.948	ug/l	97
20) Methylene Chloride	2.782	84	5812	1.126	ug/l #	84
21) trans-1,2-Dichloroethene	3.087	96	4400	0.986	ug/l	96
22) Diisopropyl ether	3.770	45	17394	1.031	ug/l #	83
23) Vinyl Acetate	3.727	43	88904	4.703	ug/l	100
24) 1,1-Dichloroethane	3.599	63	9068	1.008	ug/l	95
25) 2-Butanone	4.593	43	18933	4.665	ug/l	97
26) 2,2-Dichloropropane	4.471	77	7961	1.046	ug/l	96
27) cis-1,2-Dichloroethene	4.489	96	5670	1.016	ug/l	95
28) Bromochloromethane	4.891	49	4636	1.131	ug/l #	92
29) Tetrahydrofuran	5.025	42	12230	4.547	ug/l #	92
30) Chloroform	5.093	83	9237	1.022	ug/l	99
31) Cyclohexane	5.464	56	7483	0.989	ug/l	91
32) 1,1,1-Trichloroethane	5.385	97	7582	0.991	ug/l #	51
36) 1,1-Dichloropropene	5.696	75	5618	0.963	ug/l #	90
37) Ethyl Acetate	4.733	43	5298	0.640	ug/l #	81
38) Carbon Tetrachloride	5.672	117	5256	0.863	ug/l	83
39) Methylcyclohexane	7.373	83	8251	1.073	ug/l	93
40) Benzene	6.037	78	18885	0.995	ug/l #	87

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082224\
 Data File : VX043084.D
 Acq On : 22 Aug 2024 12:37
 Operator : JC/MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Aug 23 04:18:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082224W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 04:08:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.946	41	3041	0.759	ug/l #	70
42) 1,2-Dichloroethane	6.092	62	6294	0.925	ug/l	97
43) Isopropyl Acetate	6.354	43	12801	0.982	ug/l	98
44) Trichloroethene	7.135	130	4423	1.022	ug/l	97
45) 1,2-Dichloropropane	7.427	63	4546	0.950	ug/l #	84
46) Dibromomethane	7.592	93	3713	1.059	ug/l	96
47) Bromodichloromethane	7.830	83	6616	0.956	ug/l #	96
48) Methyl methacrylate	7.708	41	4834	0.837	ug/l #	78
49) 1,4-Dioxane	7.684	88	1866	16.021	ug/l #	95
51) 4-Methyl-2-Pentanone	8.580	43	36201	4.619	ug/l	97
52) Toluene	8.720	92	10876	0.946	ug/l	85
53) t-1,3-Dichloropropene	8.988	75	5922	0.843	ug/l	97
54) cis-1,3-Dichloropropene	8.372	75	7448	0.965	ug/l #	87
55) 1,1,2-Trichloroethane	9.153	97	4454	0.958	ug/l	93
56) Ethyl methacrylate	9.122	69	7316	0.913	ug/l	94
57) 1,3-Dichloropropane	9.311	76	8047	0.992	ug/l	95
58) 2-Chloroethyl Vinyl ether	8.251	63	17899	4.675	ug/l	99
59) 2-Hexanone	9.439	43	27466	4.559	ug/l	100
60) Dibromochloromethane	9.525	129	4524	0.882	ug/l	96
61) 1,2-Dibromoethane	9.616	107	4133	0.897	ug/l	99
64) Tetrachloroethene	9.275	164	3877	1.071	ug/l	89
65) Chlorobenzene	10.079	112	12666	1.049	ug/l	97
66) 1,1,1,2-Tetrachloroethane	10.165	131	4256	1.024	ug/l #	65
67) Ethyl Benzene	10.195	91	20174	0.971	ug/l	95
68) m/p-Xylenes	10.305	106	15206	1.923	ug/l	99
69) o-Xylene	10.646	106	8106	1.010	ug/l	93
70) Styrene	10.659	104	11919	0.908	ug/l	97
71) Bromoform	10.805	173	2896	0.844	ug/l #	98
73) Isopropylbenzene	10.963	105	18585	1.028	ug/l	97
74) N-amyl acetate	10.848	43	8882	0.926	ug/l #	93
75) 1,1,2,2-Tetrachloroethane	11.213	83	7010	1.044	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	7718	1.432	ug/l	100
77) Bromobenzene	11.201	156	4036	0.987	ug/l	91
78) n-propylbenzene	11.305	91	19941	0.980	ug/l	100
79) 2-Chlorotoluene	11.366	91	14476	1.135	ug/l	94
80) 1,3,5-Trimethylbenzene	11.451	105	14971	1.018	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.024	75	2237	0.962	ug/l	95
82) 4-Chlorotoluene	11.457	91	14697	1.032	ug/l	99
83) tert-Butylbenzene	11.713	119	15810	1.033	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	13695	0.929	ug/l	91
85) sec-Butylbenzene	11.890	105	18181	0.982	ug/l	96
86) p-Isopropyltoluene	12.012	119	14235	0.947	ug/l	93
87) 1,3-Dichlorobenzene	11.975	146	7978	1.035	ug/l	99
88) 1,4-Dichlorobenzene	11.975	146	7978	1.030	ug/l	98
89) n-Butylbenzene	12.335	91	11913	0.899	ug/l	94
90) Hexachloroethane	12.536	117	3209	1.028	ug/l	93
91) 1,2-Dichlorobenzene	12.335	146	7565	0.975	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	12.945	75	1266	0.867	ug/l	91
93) 1,2,4-Trichlorobenzene	13.591	180	3721	0.882	ug/l	95
94) Hexachlorobutadiene	13.725	225	1660	1.011	ug/l	93
95) Naphthalene	13.780	128	14739	0.894	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	4334	0.979	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082224\
Data File : VX043084.D
Acq On : 22 Aug 2024 12:37
Operator : JC/MD
Sample : VSTDIC001
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :


Quant Time: Aug 23 04:18:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082224W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 04:08:41 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 23 04:18:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082224W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 04:08:41 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:168 Resp: 369350
Ion Ratio Lower Upper
168 100
99 65.7 56.6 85.0



#2
Dichlorodifluoromethane
Concen: 0.881 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Tgt Ion: 85 Resp: 2910
Ion Ratio Lower Upper
85 100
87 37.9 16.7 50.1









#3
Chloromethane
Concen: 1.142 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 50 Resp: 5657
Ion Ratio Lower Upper
50 100
52 32.8 25.5 38.3



#4
Vinyl Chloride
Concen: 1.076 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Tgt Ion: 62 Resp: 5644
Ion Ratio Lower Upper
62 100
64 38.7 25.4 38.0#









#5
 Bromomethane
 Concen: 1.013 ug/l
 RT: 1.599 min Scan# 84
 Delta R.T. 0.000 min
 Lab File: VX043084.D
 Acq: 22 Aug 2024 12:37




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 94 Resp: 3168
 Ion Ratio Lower Upper
 94 100
 96 102.5 74.6 111.8



#6
 Chloroethane
 Concen: 1.653 ug/l
 RT: 1.666 min Scan# 95
 Delta R.T. 0.000 min
 Lab File: VX043084.D
 Acq: 22 Aug 2024 12:37

Tgt Ion: 64 Resp: 4072
 Ion Ratio Lower Upper
 64 100
 66 29.7 25.1 37.7









#7
 Trichlorofluoromethane
 Concen: 1.081 ug/l
 RT: 1.867 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX043084.D
 Acq: 22 Aug 2024 12:37




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:101 Resp: 9583
 Ion Ratio Lower Upper
 101 100
 103 57.2 51.4 77.2



#8
 Diethyl Ether
 Concen: 1.060 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. 0.000 min
 Lab File: VX043084.D
 Acq: 22 Aug 2024 12:37

Tgt Ion: 74 Resp: 4064
 Ion Ratio Lower Upper
 74 100
 45 99.8 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 1.010 ug/l
RT: 2.313 min Scan# 202
Delta R.T. -0.006 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 4363
Ion Ratio Lower Upper
101 100
85 58.4 37.0 55.6#
151 73.0 58.6 87.8



#10
Methyl Iodide
Concen: 1.012 ug/l
RT: 2.441 min Scan# 222
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Tgt Ion:142 Resp: 4632
Ion Ratio Lower Upper
142 100
127 43.1 41.6 62.4
141 12.8 11.6 17.4





#11

Tert butyl alcohol

Concen: 4.950 ug/l

RT: 2.995 min Scan# 313

Delta R.T. 0.000 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 59 Resp: 6225

Ion Ratio Lower Upper

59 100

57 10.7 6.0 9.0#



#12

1,1-Dichloroethene

Concen: 1.048 ug/l

RT: 2.306 min Scan# 200

Delta R.T. 0.000 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37




Tgt Ion: 96 Resp: 4548


Ion Ratio Lower Upper

96 100

61 174.7 144.6 216.8

98 58.9 50.4 75.6





#13

Acrolein

Concen: Below Cal

RT: 2.239 min Scan# 189

Delta R.T. 0.000 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 56 Resp: 4819

Ion Ratio Lower Upper

56 100

55 69.4 57.1 85.7



#14

Allyl chloride

Concen: 1.070 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.000 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37




Tgt Ion: 41 Resp: 8239


Ion Ratio Lower Upper

41 100

39 68.3 55.8 83.6

76 34.7 24.3 36.5









#15
Acrylonitrile
Concen: 4.914 ug/l
RT: 3.075 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 53 Resp: 14507
Ion Ratio Lower Upper
53 100
52 82.8 66.0 99.0
51 37.3 29.8 44.6



#16
Acetone
Concen: 5.621 ug/l
RT: 2.392 min Scan# 214
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37


Tgt Ion: 43 Resp: 14977
Ion Ratio Lower Upper
43 100
58 28.6 21.9 32.9








#17
Carbon Disulfide
Concen: 1.066 ug/l
RT: 2.502 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 76 Resp: 11923
Ion Ratio Lower Upper
76 100
78 12.1 7.0 10.6#



#18
Methyl Acetate
Concen: 1.075 ug/l
RT: 2.709 min Scan# 266
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Tgt Ion: 43 Resp: 9088
Ion Ratio Lower Upper
43 100
74 21.0 17.8 26.8









#19
Methyl tert-butyl Ether
Concen: 0.948 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 15481
Ion Ratio Lower Upper
73 100
57 22.9 17.3 25.9



#20
Methylene Chloride
Concen: 1.126 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Tgt Ion: 84 Resp: 5812
Ion Ratio Lower Upper
84 100
49 153.0 103.6 155.4
51 30.7 33.8 50.6#
86 56.9 49.4 74.0









#21
trans-1,2-Dichloroethene
Concen: 0.986 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 96 Resp: 4400
Ion Ratio Lower Upper
96 100
61 145.0 119.2 178.8
98 66.3 48.7 73.1



#22
Diisopropyl ether
Concen: 1.031 ug/l
RT: 3.770 min Scan# 440
Delta R.T. 0.006 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Tgt Ion: 45 Resp: 17394
Ion Ratio Lower Upper
45 100
43 106.6 69.7 104.5#
87 21.1 20.8 31.2
59 12.7 9.0 13.4









#23
 Vinyl Acetate
 Concen: 4.703 ug/l
 RT: 3.727 min Scan# 41
 Delta R.T. 0.006 min
 Lab File: VX043084.D
 Acq: 22 Aug 2024 12:37




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 88904
 Ion Ratio Lower Upper
 43 100
 86 9.9 7.8 11.8



#24
 1,1-Dichloroethane
 Concen: 1.008 ug/l
 RT: 3.599 min Scan# 412
 Delta R.T. -0.006 min
 Lab File: VX043084.D
 Acq: 22 Aug 2024 12:37

Tgt Ion: 63 Resp: 9068
 Ion Ratio Lower Upper
 63 100
 98 4.6 3.3 9.8
 100 5.6 2.1 6.5









#25
2-Butanone
Concen: 4.665 ug/l
RT: 4.593 min Scan# 51
Delta R.T. 0.031 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 18933
Ion Ratio Lower Upper
43 100
72 25.7 19.4 29.2



#26
2,2-Dichloropropane
Concen: 1.046 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Tgt Ion: 77 Resp: 7961
Ion Ratio Lower Upper
77 100
97 19.7 10.8 32.3





#27

cis-1,2-Dichloroethene

Concen: 1.016 ug/l

RT: 4.489 min Scan# 557

Delta R.T. 0.006 min


Lab File: VX043084.D

Acq: 22 Aug 2024 12:37

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 96 Resp: 5670

Ion Ratio Lower Upper

96 100

61 147.4 0.0 312.2

98 66.0 0.0 129.0



#28

Bromochloromethane

Concen: 1.131 ug/l

RT: 4.891 min Scan# 624

Delta R.T. 0.000 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37




Tgt Ion: 49 Resp: 4636


Ion Ratio Lower Upper

49 100

129 0.0 0.0 3.8

130 65.3 57.5 86.3





#29

Tetrahydrofuran

Concen: 4.547 ug/l

RT: 5.025 min Scan# 64

Delta R.T. 0.018 min


Lab File: VX043084.D

Acq: 22 Aug 2024 12:37




Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion: 42	Resp: 12230
Ion Ratio	Lower Upper
42	100
72	44.4 34.2 51.4
71	49.3 32.2 48.2#



#30

Chloroform

Concen: 1.022 ug/l




RT: 5.093 min Scan# 657


Delta R.T. 0.006 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37


Tgt Ion: 83	Resp: 9237
Ion Ratio	Lower Upper
83	100
85	63.4 50.3 75.5








#31
Cyclohexane
Concen: 0.989 ug/l
RT: 5.464 min Scan# 718
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 56 Resp: 7483
Ion Ratio Lower Upper
56 100
69 34.3 23.4 35.2
84 74.3 65.9 98.9




#32
1,1,1-Trichloroethane
Concen: 0.991 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37



Tgt Ion: 97 Resp: 7582
Ion Ratio Lower Upper
97 100
99 0.0 51.1 76.7#
61 47.9 36.4 54.6









#33
 1,2-Dichloroethane-d4
 Concen: 1.259 ug/l
 RT: 5.964 min Scan# 80
 Delta R.T. 0.012 min
 Lab File: VX043084.D
 Acq: 22 Aug 2024 12:37




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 65 Resp: 7641
 Ion Ratio Lower Upper
 65 100
 67 59.1 0.0 100.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.757 min Scan# 930
 Delta R.T. 0.000 min
 Lab File: VX043084.D
 Acq: 22 Aug 2024 12:37


Tgt Ion: 114 Resp: 667190
 Ion Ratio Lower Upper
 114 100
 63 21.1 0.0 45.0
 88 15.7 0.0 36.6








#35
Dibromofluoromethane
Concen: 1.266 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:113 Resp: 5829
Ion Ratio Lower Upper
113 100
111 98.0 82.9 124.3
192 18.1 13.3 19.9



#36
1,1-Dichloropropene
Concen: 0.963 ug/l
RT: 5.696 min Scan# 756
Delta R.T. 0.012 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Tgt Ion: 75 Resp: 5618
Ion Ratio Lower Upper
75 100
110 36.8 17.0 50.9
77 39.6 24.7 37.1#









#37
Ethyl Acetate
Concen: 0.640 ug/l
RT: 4.733 min Scan# 598
Delta R.T. 0.012 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 5298
Ion Ratio Lower Upper
43 100
61 0.0 10.6 16.0#
70 10.9 8.3 12.5



#38
Carbon Tetrachloride
Concen: 0.863 ug/l
RT: 5.672 min Scan# 752
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37


Tgt Ion: 117 Resp: 5256
Ion Ratio Lower Upper
117 100
119 117.1 78.2 117.2
121 34.0 24.4 36.6








#39
Methylcyclohexane
Concen: 1.073 ug/l
RT: 7.373 min Scan# 1031
Delta R.T. -0.006 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 83 Resp: 8251
Ion Ratio Lower Upper
83 100
55 71.8 64.6 97.0
98 44.8 36.7 55.1



#40
Benzene
Concen: 0.995 ug/l
RT: 6.037 min Scan# 812
Delta R.T. 0.006 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37


Tgt Ion: 78 Resp: 18885
Ion Ratio Lower Upper
78 100
77 17.5 19.3 28.9#








#41
Methacrylonitrile
Concen: 0.759 ug/l
RT: 4.946 min Scan# 61
Delta R.T. 0.025 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:	41	Resp:	3041
Ion	Ratio	Lower	Upper
41	100		
39	37.4	45.2	67.8#
67	100.4	52.6	78.8#
52	25.5	22.7	34.1



#42
1,2-Dichloroethane
Concen: 0.925 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.012 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Tgt Ion:	62	Resp:	6294
Ion	Ratio	Lower	Upper
62	100		
98	9.2	0.0	16.0





#43

Isopropyl Acetate

Concen: 0.982 ug/l

RT: 6.354 min Scan# 862

Delta R.T. 0.012 min


Lab File: VX043084.D

Acq: 22 Aug 2024 12:37




Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion: 43	Resp: 12801
Ion Ratio	Lower Upper
43 100	
61 20.3	15.2 22.8
87 11.5	9.8 14.6



#44

Trichloroethene

Concen: 1.022 ug/l




RT: 7.135 min Scan# 992


Delta R.T. 0.012 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37

Tgt Ion:130	Resp:	4423	
Ion	Ratio	Lower	Upper
130	100		
95	99.2	0.0	204.4





#45

1,2-Dichloropropane

Concen: 0.950 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 63 Resp: 4546

Ion Ratio Lower Upper

63 100

65 38.9 24.3 36.5#



#46

Dibromomethane

Concen: 1.059 ug/l

RT: 7.592 min Scan# 1067

Delta R.T. 0.012 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37




Tgt Ion: 93 Resp: 3713


Ion Ratio Lower Upper

93 100

95 83.8 66.6 100.0


174 79.3 69.4 104.0








#47
Bromodichloromethane
Concen: 0.956 ug/l
RT: 7.830 min Scan# 1106
Delta R.T. 0.012 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 83 Resp: 6616
Ion Ratio Lower Upper
83 100
85 65.2 49.8 74.6
127 11.1 7.1 10.7#



#48
Methyl methacrylate
Concen: 0.837 ug/l
RT: 7.708 min Scan# 1086
Delta R.T. 0.012 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37


Tgt Ion: 41 Resp: 4834
Ion Ratio Lower Upper
41 100
69 99.3 59.2 88.8#
39 49.6 47.7 71.5








#49
1,4-Dioxane
Concen: 16.021 ug/l
RT: 7.684 min Scan# 1082
Delta R.T. 0.018 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 88 Resp: 1866
Ion Ratio Lower Upper
88 100
43 37.8 0.0 0.0#
58 70.9 60.1 90.1



#50
Toluene-d8
Concen: 1.126 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Tgt Ion: 98 Resp: 18698
Ion Ratio Lower Upper
98 100
100 61.7 53.1 79.7





#51

4-Methyl-2-Pentanone

Concen: 4.619 ug/l

RT: 8.580 min Scan# 1229

Delta R.T. 0.006 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 36201

Ion Ratio Lower Upper

43 100

58 39.8 30.4 45.6



#52

Toluene

Concen: 0.946 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. 0.006 min

Lab File: VX043084.D




Acq: 22 Aug 2024 12:37


Tgt Ion: 92 Resp: 10876

Ion Ratio Lower Upper

92 100

91 188.2 134.2 201.2









#53
 t-1,3-Dichloropropene
 Concen: 0.843 ug/l
 RT: 8.988 min Scan# 1195
 Delta R.T. 0.012 min
 Lab File: VX043084.D
 Acq: 22 Aug 2024 12:37




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 75 Resp: 5922
 Ion Ratio Lower Upper
 75 100
 77 30.3 25.7 38.5



#54
 cis-1,3-Dichloropropene
 Concen: 0.965 ug/l
 RT: 8.372 min Scan# 1195
 Delta R.T. 0.006 min
 Lab File: VX043084.D
 Acq: 22 Aug 2024 12:37


Tgt Ion: 75 Resp: 7448
 Ion Ratio Lower Upper
 75 100
 77 23.2 25.3 37.9#
 39 49.2 46.7 70.1








#55
 1,1,2-Trichloroethane
 Concen: 0.958 ug/l
 RT: 9.153 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX043084.D
 Acq: 22 Aug 2024 12:37

Instrument :
 MSVOA_X
 ClientSampleId :



Tgt Ion: 97 Resp: 4454




Ion	Ratio	Lower	Upper
97	100		
83	74.8	65.8	98.8
85	54.0	41.6	62.4
99	67.1	47.8	71.8




#56
 Ethyl methacrylate
 Concen: 0.913 ug/l
 RT: 9.122 min Scan# 1318
 Delta R.T. 0.006 min
 Lab File: VX043084.D
 Acq: 22 Aug 2024 12:37

Tgt Ion: 69 Resp: 7316

Ion	Ratio	Lower	Upper
69	100		
41	64.9	56.8	85.2
39	38.1	32.0	48.0





#57

1,3-Dichloropropane

Concen: 0.992 ug/l

RT: 9.311 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 8047

Ion Ratio Lower Upper

76 100

78 30.0 26.1 39.1



#58

2-Chloroethyl Vinyl ether

Concen: 4.675 ug/l

RT: 8.251 min Scan# 1175

Delta R.T. 0.012 min

Lab File: VX043084.D




Acq: 22 Aug 2024 12:37


Tgt Ion: 63 Resp: 17899

Ion Ratio Lower Upper

63 100

106 27.8 22.5 33.7





#59

2-Hexanone

Concen: 4.559 ug/l

RT: 9.439 min Scan# 1370

Delta R.T. 0.006 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 27466

Ion Ratio Lower Upper

43 100

58 52.3 26.2 78.5



#60

Dibromochloromethane

Concen: 0.882 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. 0.006 min

Lab File: VX043084.D




Acq: 22 Aug 2024 12:37


Tgt Ion: 129 Resp: 4524

Ion Ratio Lower Upper

129 100


127 80.9 38.8 116.4








#61
1,2-Dibromoethane
Concen: 0.897 ug/l
RT: 9.616 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:107 Resp: 4133
Ion Ratio Lower Upper
107 100
109 92.9 75.2 112.8



#62
4-Bromofluorobenzene
Concen: 1.003 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.006 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37


Tgt Ion: 95 Resp: 5798
Ion Ratio Lower Upper
95 100
174 73.4 0.0 135.6
176 75.5 0.0 131.0








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:117 Resp: 551005
Ion Ratio Lower Upper
117 100
82 58.1 46.2 69.4
119 33.1 25.5 38.3



#64
Tetrachloroethene
Concen: 1.071 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37


Tgt Ion:164 Resp: 3877
Ion Ratio Lower Upper
164 100
166 146.9 102.7 154.1
129 109.1 80.4 120.6
131 103.9 76.5 114.7








#65
Chlorobenzene
Concen: 1.049 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:112 Resp: 12666
Ion Ratio Lower Upper
112 100
114 30.2 25.7 38.5




#66
1,1,1,2-Tetrachloroethane
Concen: 1.024 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.006 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37



Tgt Ion:131 Resp: 4256
Ion Ratio Lower Upper
131 100
133 91.7 47.1 141.4
119 0.0 32.2 96.6#









#67
Ethyl Benzene
Concen: 0.971 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 20174
Ion Ratio Lower Upper
91 100
106 33.7 24.9 37.3



#68
m/p-Xylenes
Concen: 1.923 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.006 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Tgt Ion: 106 Resp: 15206
Ion Ratio Lower Upper
106 100
91 205.5 165.5 248.3





#69

o-Xylene

Concen: 1.010 ug/l

RT: 10.646 min Scan# 1567

Delta R.T. 0.006 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:106 Resp: 8106

Ion Ratio Lower Upper

106 100

91 206.2 108.9 326.8



#70

Styrene

Concen: 0.908 ug/l

RT: 10.659 min Scan# 1570

Delta R.T. 0.006 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37




Tgt Ion:104 Resp: 11919


Ion Ratio Lower Upper

104 100

78 50.7 43.6 65.4

103 54.2 43.6 65.4





#71

Bromoform

Concen: 0.844 ug/l

RT: 10.805 min Scan# 11

Delta R.T. 0.006 min


Lab File: VX043084.D

Acq: 22 Aug 2024 12:37

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion:173 Resp: 2896

Ion Ratio Lower Upper

173 100

175 49.6 24.1 72.4

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.000 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37




Tgt Ion:152 Resp: 231473


Ion Ratio Lower Upper

152 100

115 61.2 43.3 129.9

150 157.7 0.0 345.0





#73

Isopropylbenzene

Concen: 1.028 ug/l

RT: 10.963 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:105 Resp: 18585

Ion Ratio Lower Upper

105 100

120 27.2 12.9 38.6



#74

N-ethyl acetate

Concen: 0.926 ug/l

RT: 10.848 min Scan# 1601

Delta R.T. 0.006 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37

Tgt Ion: 43 Resp: 8882




Ion Ratio Lower Upper


43 100

70 38.8 32.9 49.3

55 29.9 19.3 28.9#

61 20.0 18.8 28.2





#75

1,1,2,2-Tetrachloroethane

Concen: 1.044 ug/l

RT: 11.213 min Scan# 1661

Delta R.T. 0.000 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 7010

Ion Ratio Lower Upper

83 100

131 10.6 5.3 15.8

85 65.4 32.0 96.2



#76

1,2,3-Trichloropropane

Concen: 1.432 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. 0.000 min

Lab File: VX043084.D




Acq: 22 Aug 2024 12:37


Tgt Ion: 75 Resp: 7718

Ion Ratio Lower Upper

75 100

77 38.5 19.1 57.5





#77

Bromobenzene

Concen: 0.987 ug/l

RT: 11.201 min Scan# 1659

Delta R.T. 0.006 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:156 Resp: 4036

Ion Ratio Lower Upper

156 100

77 186.8 88.4 265.3

158 108.1 48.0 144.0



#78

n-propylbenzene

Concen: 0.980 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. 0.000 min

Lab File: VX043084.D




Acq: 22 Aug 2024 12:37


Tgt Ion: 91 Resp: 19941

Ion Ratio Lower Upper

91 100

120 22.7 11.5 34.4





#79

2-Chlorotoluene

Concen: 1.135 ug/l

RT: 11.366 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 14476

Ion Ratio Lower Upper

91 100

126 28.9 16.3 48.8



#80

1,3,5-Trimethylbenzene

Concen: 1.018 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX043084.D




Acq: 22 Aug 2024 12:37


Tgt Ion: 105 Resp: 14971

Ion Ratio Lower Upper

105 100

120 47.1 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 0.962 ug/l
RT: 11.024 min Scan# 1630
Delta R.T. 0.006 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 2237
Ion Ratio Lower Upper
75 100
53 119.8 90.5 135.7
89 48.4 36.9 55.3



#82
4-Chlorotoluene
Concen: 1.032 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Tgt Ion: 91 Resp: 14697
Ion Ratio Lower Upper
91 100
126 28.0 14.2 42.6









#83
tert-Butylbenzene
Concen: 1.033 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
119	100		
91	60.3	31.2	93.6
134	25.4	11.7	35.0



#84
1,2,4-Trimethylbenzene
Concen: 0.929 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.006 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Tgt Ion	Ratio	Lower	Upper
105	100		
120	47.9	21.2	63.6





#85

sec-Butylbenzene

Concen: 0.982 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. 0.000 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:105 Resp: 18181

Ion Ratio Lower Upper

105 100

134 17.8 9.8 29.4



#86

p-Isopropyltoluene

Concen: 0.947 ug/l

RT: 12.012 min Scan# 1792

Delta R.T. 0.006 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37




Tgt Ion:119 Resp: 14235


Ion Ratio Lower Upper

119 100

134 27.6 12.8 38.3


91 30.9 13.1 39.3







#87
1,3-Dichlorobenzene
Concen: 1.035 ug/l
RT: 11.975 min Scan# 1786
Delta R.T. 0.006 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:146 Resp: 7978
Ion Ratio Lower Upper
146 100
111 45.6 22.1 66.1
148 63.3 31.8 95.3



#88
1,4-Dichlorobenzene
Concen: 1.030 ug/l
RT: 11.975 min Scan# 1786
Delta R.T. -0.067 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Tgt Ion:146 Resp: 7978
Ion Ratio Lower Upper
146 100
111 45.6 21.3 63.7
148 63.3 31.8 95.4





#89

n-Butylbenzene

Concen: 0.899 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.006 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 91 Resp: 11913

Ion Ratio Lower Upper

91 100

92 51.4 27.4 82.0

134 21.3 12.6 37.8



#90

Hexachloroethane

Concen: 1.028 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX043084.D




Acq: 22 Aug 2024 12:37


Tgt Ion: 117 Resp: 3209

Ion Ratio Lower Upper

117 100

201 60.1 32.9 98.6





#91

1,2-Dichlorobenzene

Concen: 0.975 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 7565

Ion Ratio Lower Upper

146 100

111 46.6 22.4 67.0

148 60.2 31.6 95.0



#92

1,2-Dibromo-3-Chloropropane

Concen: 0.867 ug/l

RT: 12.945 min Scan# 1945

Delta R.T. 0.006 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37




Tgt Ion: 75 Resp: 1266


Ion Ratio Lower Upper

75 100

155 73.8 36.1 108.5

157 103.1 44.9 134.7





#93

1,2,4-Trichlorobenzene

Concen: 0.882 ug/l

RT: 13.591 min Scan# 2050

Delta R.T. 0.006 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:180 Resp: 3721

Ion Ratio Lower Upper

180 100

182 94.6 48.3 144.8

145 27.8 18.0 54.0



#94

Hexachlorobutadiene

Concen: 1.011 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX043084.D

Acq: 22 Aug 2024 12:37




Tgt Ion:225 Resp: 1660


Ion Ratio Lower Upper

225 100

223 56.6 31.1 93.2

227 70.4 32.5 97.5









#95
Naphthalene
Concen: 0.894 ug/l
RT: 13.780 min Scan# 2081
Delta R.T. 0.006 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:128	Resp:	14739
Ion Ratio	Lower	Upper
128	100	
127	13.3	10.3 15.5
129	10.9	8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 0.979 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX043084.D
Acq: 22 Aug 2024 12:37

Tgt Ion:180	Resp:	4334
Ion Ratio	Lower	Upper
180	100	
182	91.8	46.9 140.6
145	33.1	17.8 53.5

