

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
 Data File : VX030721.D
 Acq On : 23 Aug 2022 10:45
 Operator : JC/MD
 Sample : VX0823MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0823MBL01

Quant Time: Aug 24 00:45:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	138270	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	307769	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	282231	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	120265	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	184439	66.365	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	132.740%#
35) Dibromofluoromethane	5.385	113	171120	58.162	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	116.320%
50) Toluene-d8	8.653	98	625024	54.857	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.720%
62) 4-Bromofluorobenzene	11.079	95	220322	52.104	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
Data File : VX030721.D
Acq On : 23 Aug 2022 10:45
Operator : JC/MD
Sample : VX0823MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0823MBL01

Quant Time: Aug 24 00:45:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

