

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
 Data File : VX030728.D
 Acq On : 23 Aug 2022 13:44
 Operator : JC/MD
 Sample : N4245-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW178S-20220816

Quant Time: Aug 24 00:49:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	128869	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	285904	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	257918	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108006	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	150398	58.065	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.120%
35) Dibromofluoromethane	5.397	113	138872	50.811	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.620%
50) Toluene-d8	8.653	98	498371	47.086	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.180%
62) 4-Bromofluorobenzene	11.085	95	169707	43.203	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.400%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1443	1.088	ug/l	94
19) Methyl tert-butyl Ether	3.123	73	10054	1.458	ug/l	95
44) Trichloroethene	7.123	130	5567	1.646	ug/l	99
64) Tetrachloroethene	9.281	164	1001	0.361	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
Data File : VX030728.D
Acq On : 23 Aug 2022 13:44
Operator : JC/MD
Sample : N4245-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW178S-20220816

Quant Time: Aug 24 00:49:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

