

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
 Data File : VX030731.D
 Acq On : 23 Aug 2022 14:55
 Operator : JC/MD
 Sample : N4167-01DL 10X
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE137-CARBON-20220810DL

Quant Time: Aug 24 00:50:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	130005	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	285796	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	251621	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92250	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	149572	57.241	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.480%
35) Dibromofluoromethane	5.391	113	135409	49.563	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.120%
50) Toluene-d8	8.653	98	504431	47.676	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.360%
62) 4-Bromofluorobenzene	11.079	95	162128	41.289	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.580%#
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	101240	48.726	ug/l	99
30) Chloroform	5.105	83	4905	1.176	ug/l #	82
38) Carbon Tetrachloride	5.678	117	20302	5.531	ug/l	94
44) Trichloroethene	7.129	130	11113	3.288	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
Data File : VX030731.D
Acq On : 23 Aug 2022 14:55
Operator : JC/MD
Sample : N4167-01DL 10X
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE137-CARBON-20220810DL

Quant Time: Aug 24 00:50:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

