

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
 Data File : VX030732.D
 Acq On : 23 Aug 2022 15:18
 Operator : JC/MD
 Sample : N4245-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-20220816

Quant Time: Aug 24 00:50:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	125223	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	281231	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	248761	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92895	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	152831	60.722	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	= 121.440%		
35) Dibromofluoromethane	5.391	113	145045	53.952	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 107.900%		
50) Toluene-d8	8.653	98	520404	49.984	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	= 99.960%		
62) 4-Bromofluorobenzene	11.085	95	172627	44.677	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	= 89.360%		
Target Compounds						Qvalue
16) Acetone	2.386	43	1972	1.530	ug/l	92
44) Trichloroethene	7.135	130	1358	0.408	ug/l	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
Data File : VX030732.D
Acq On : 23 Aug 2022 15:18
Operator : JC/MD
Sample : N4245-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-20220816

Quant Time: Aug 24 00:50:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

