

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
 Data File : VX030733.D
 Acq On : 23 Aug 2022 15:42
 Operator : JC/MD
 Sample : N4245-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW179D-20220816

Quant Time: Aug 24 00:51:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	130702	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	290714	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	262559	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108160	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	155833	59.319	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 118.640%	
35) Dibromofluoromethane	5.391	113	146126	52.581	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.160%	
50) Toluene-d8	8.652	98	527989	49.059	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.120%	
62) 4-Bromofluorobenzene	11.085	95	183856	46.031	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.060%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.318	96	878	0.430	ug/l	92
16) Acetone	2.391	43	1863	1.385	ug/l #	80
19) Methyl tert-butyl Ether	3.117	73	6522	0.933	ug/l #	84
27) cis-1,2-Dichloroethene	4.495	96	71862	27.116	ug/l	91
44) Trichloroethene	7.128	130	304769	88.639	ug/l	100
64) Tetrachloroethene	9.274	164	27641	9.779	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
Data File : VX030733.D
Acq On : 23 Aug 2022 15:42
Operator : JC/MD
Sample : N4245-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW179D-20220816

Quant Time: Aug 24 00:51:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

