

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
 Data File : VX030734.D
 Acq On : 23 Aug 2022 16:05
 Operator : JC/MD
 Sample : N4245-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW179D1-20220816

Quant Time: Aug 24 00:51:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	128471	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	290139	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	261124	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107765	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	152767	59.162	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 118.320%			
35) Dibromofluoromethane	5.391	113	143169	51.619	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.240%			
50) Toluene-d8	8.653	98	514457	47.896	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 95.800%			
62) 4-Bromofluorobenzene	11.085	95	173844	43.610	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 87.220%			
Target Compounds					Qvalue	
19) Methyl tert-butyl Ether	3.117	73	3115	0.453	ug/l	94
27) cis-1,2-Dichloroethene	4.495	96	35155	13.496	ug/l	90
44) Trichloroethene	7.129	130	281082	81.912	ug/l	97
64) Tetrachloroethene	9.275	164	26425	9.400	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
Data File : VX030734.D
Acq On : 23 Aug 2022 16:05
Operator : JC/MD
Sample : N4245-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW179D1-20220816

Quant Time: Aug 24 00:51:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

