

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
 Data File : VX030735.D
 Acq On : 23 Aug 2022 16:28
 Operator : JC/MD
 Sample : N4245-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20220816

Quant Time: Aug 24 00:51:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	123935	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	276298	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	245535	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96543	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	152087	61.054	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	122.100%
35) Dibromofluoromethane	5.391	113	142233	53.850	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.700%
50) Toluene-d8	8.653	98	517112	50.555	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.120%
62) 4-Bromofluorobenzene	11.085	95	173075	45.592	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.180%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1960	1.537	ug/l	92
19) Methyl tert-butyl Ether	3.117	73	6352	0.958	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	71205	28.335	ug/l	90
44) Trichloroethene	7.129	130	304483	93.176	ug/l	99
64) Tetrachloroethene	9.281	164	27329	10.339	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
Data File : VX030735.D
Acq On : 23 Aug 2022 16:28
Operator : JC/MD
Sample : N4245-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP02-20220816

Quant Time: Aug 24 00:51:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

