

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
 Data File : VX030737.D
 Acq On : 23 Aug 2022 17:15
 Operator : JC/MD
 Sample : N4245-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-MW01D1-20220816

Quant Time: Aug 24 00:52:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	125938	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	275790	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	253820	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99270	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	155479	61.423	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	122.840%
35) Dibromofluoromethane	5.391	113	145305	55.115	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.220%
50) Toluene-d8	8.653	98	527212	51.637	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.280%
62) 4-Bromofluorobenzene	11.085	95	175171	46.230	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.460%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	2008	1.549	ug/l	92
29) Tetrahydrofuran	5.019	42	15261	10.415	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
Data File : VX030737.D
Acq On : 23 Aug 2022 17:15
Operator : JC/MD
Sample : N4245-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D1-20220816

Quant Time: Aug 24 00:52:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

