

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
 Data File : VX030738.D
 Acq On : 23 Aug 2022 17:38
 Operator : JC/MD
 Sample : N4245-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE117D2-20220816

Quant Time: Aug 24 00:53:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	122313	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	272559	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236333	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84149	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	150326	61.147	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	122.300%
35) Dibromofluoromethane	5.385	113	140087	53.765	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.540%
50) Toluene-d8	8.653	98	503628	49.912	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.820%
62) 4-Bromofluorobenzene	11.085	95	161953	43.248	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
Data File : VX030738.D
Acq On : 23 Aug 2022 17:38
Operator : JC/MD
Sample : N4245-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE117D2-20220816

Quant Time: Aug 24 00:53:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

