

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
 Data File : VX030742.D
 Acq On : 23 Aug 2022 19:13
 Operator : JC/MD
 Sample : N4245-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE126D2-20220817

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 08/24/2022
 Supervised By :Mahesh Dadoda 08/24/2022

Quant Time: Aug 24 00:54:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	123411	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	275304	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	250513	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98389	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	153751	61.984	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 123.960%			
35) Dibromofluoromethane	5.391	113	144712	54.987	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 109.980%			
50) Toluene-d8	8.653	98	517758	50.801	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.600%			
62) 4-Bromofluorobenzene	11.085	95	175011	46.269	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 92.540%			
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.318	96	2214	1.148	ug/l #	79
16) Acetone	2.385	43	1521	1.198	ug/l	90
27) cis-1,2-Dichloroethene	4.501	96	2353	0.940	ug/l	88
30) Chloroform	5.098	83	1578m	0.398	ug/l	
44) Trichloroethene	7.128	130	842755	258.826	ug/l	96
64) Tetrachloroethene	9.287	164	1351	0.501	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
Data File : VX030742.D
Acq On : 23 Aug 2022 19:13
Operator : JC/MD
Sample : N4245-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE126D2-20220817

Quant Time: Aug 24 00:54:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 08/24/2022
Supervised By :Mahesh Dadoda 08/24/2022

