

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
 Data File : VX030746.D
 Acq On : 23 Aug 2022 21:10
 Operator : JC/MD
 Sample : BLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BLK

Quant Time: Aug 24 04:30:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	132397	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	293982	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	265852	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107005	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	163867	61.579	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	123.160%
35) Dibromofluoromethane	5.391	113	153432	54.596	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.200%
50) Toluene-d8	8.653	98	562495	51.684	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.360%
62) 4-Bromofluorobenzene	11.085	95	185585	45.947	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
Data File : VX030746.D
Acq On : 23 Aug 2022 21:10
Operator : JC/MD
Sample : BLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BLK

Quant Time: Aug 24 04:30:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

