

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
 Data File : VX030747.D
 Acq On : 23 Aug 2022 21:33
 Operator : JC/MD
 Sample : VX0823WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0823WBL02

Quant Time: Aug 24 04:31:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126176	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	232200	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230223	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	131539	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	130781	51.569	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	= 103.140%		
35) Dibromofluoromethane	5.391	113	113251	51.021	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 102.040%		
50) Toluene-d8	8.653	98	392348	45.642	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	= 91.280%		
62) 4-Bromofluorobenzene	11.085	95	139536	43.738	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	= 87.480%		
Target Compounds						Qvalue
44) Trichloroethene	7.129	130	1017	0.370	ug/l	58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
Data File : VX030747.D
Acq On : 23 Aug 2022 21:33
Operator : JC/MD
Sample : VX0823WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0823WBL02

Quant Time: Aug 24 04:31:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

