

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
 Data File : VX030751.D
 Acq On : 23 Aug 2022 23:07
 Operator : JC/MD
 Sample : N4245-03DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW179D2-20220816DL

Quant Time: Aug 24 04:33:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	133657	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	297374	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	267943	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99447	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	164877	61.374	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	122.740%
35) Dibromofluoromethane	5.391	113	152469	53.634	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.260%
50) Toluene-d8	8.653	98	569015	51.687	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.380%
62) 4-Bromofluorobenzene	11.085	95	182524	44.674	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.340%
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.325	96	1612	0.772	ug/l	# 80
27) cis-1,2-Dichloroethene	4.495	96	16216	5.984	ug/l	86
44) Trichloroethene	7.129	130	131604	37.418	ug/l	95
64) Tetrachloroethene	9.275	164	9769	3.387	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082322\
Data File : VX030751.D
Acq On : 23 Aug 2022 23:07
Operator : JC/MD
Sample : N4245-03DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW179D2-20220816DL

Quant Time: Aug 24 04:33:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

