

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX082322\  
 Data File : VX030723.D  
 Acq On : 23 Aug 2022 11:44  
 Operator : JC/MD  
 Sample : VX0823WBS01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0823WBS01

Manual Integrations  
 APPROVED

Reviewed By :Krupa Patel 08/24/2022  
 Supervised By :Mahesh Dadoda 08/24/2022

Quant Time: Aug 24 00:45:43 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072522W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Jul 25 16:49:17 2022  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc     | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|----------|--------|----------|
| -----                        |                |      |            |          |        |          |
| Internal Standards           |                |      |            |          |        |          |
| 1) Pentafluorobenzene        | 5.556          | 168  | 136132     | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.769          | 114  | 232560     | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.055         | 117  | 234839     | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024         | 152  | 150255     | 50.000   | ug/l   | 0.00     |
|                              |                |      |            |          |        |          |
| System Monitoring Compounds  |                |      |            |          |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.958          | 65   | 131669     | 48.122   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 74 - 125 |      | Recovery = | 96.240%  |        |          |
| 35) Dibromofluoromethane     | 5.391          | 113  | 119044     | 53.547   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 75 - 124 |      | Recovery = | 107.100% |        |          |
| 50) Toluene-d8               | 8.653          | 98   | 433872     | 50.395   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 86 - 113 |      | Recovery = | 100.780% |        |          |
| 62) 4-Bromofluorobenzene     | 11.085         | 95   | 159751     | 49.997   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 83 - 123 |      | Recovery = | 100.000% |        |          |
|                              |                |      |            |          |        |          |
| Target Compounds             |                |      |            |          | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.166          | 85   | 46450      | 20.794   | ug/l   | 97       |
| 3) Chloromethane             | 1.288          | 50   | 42496      | 19.238   | ug/l   | 98       |
| 4) Vinyl Chloride            | 1.374          | 62   | 45441      | 20.382   | ug/l   | 96       |
| 5) Bromomethane              | 1.611          | 94   | 15459      | 15.834   | ug/l   | 97       |
| 6) Chloroethane              | 1.684          | 64   | 28176      | 22.010   | ug/l   | 99       |
| 7) Trichlorofluoromethane    | 1.886          | 101  | 66854      | 22.760   | ug/l   | 95       |
| 8) Diethyl Ether             | 2.136          | 74   | 27224      | 23.668   | ug/l   | 95       |
| 9) 1,1,2-Trichlorotrifluo... | 2.325          | 101  | 45355      | 20.847   | ug/l   | 99       |
| 10) Methyl Iodide            | 2.453          | 142  | 26260      | 13.532   | ug/l   | 94       |
| 11) Tert butyl alcohol       | 2.983          | 59   | 65032      | 97.436   | ug/l   | 99       |
| 12) 1,1-Dichloroethene       | 2.318          | 96   | 40543      | 19.065   | ug/l   | 97       |
| 13) Acrolein                 | 2.239          | 56   | 3120       | 19.019   | ug/l   | 94       |
| 14) Allyl chloride           | 2.666          | 41   | 72556      | 16.534   | ug/l   | 96       |
| 15) Acrylonitrile            | 3.068          | 53   | 145647     | 88.305   | ug/l   | 98       |
| 16) Acetone                  | 2.386          | 43   | 111961     | 79.924   | ug/l   | 97       |
| 17) Carbon Disulfide         | 2.507          | 76   | 99195      | 17.280   | ug/l   | 98       |
| 18) Methyl Acetate           | 2.709          | 43   | 73594      | 17.762   | ug/l   | 99       |
| 19) Methyl tert-butyl Ether  | 3.117          | 73   | 146042     | 20.048   | ug/l   | 97       |
| 20) Methylene Chloride       | 2.788          | 84   | 51842      | 20.073   | ug/l   | 95       |
| 21) trans-1,2-Dichloroethene | 3.093          | 96   | 44321      | 18.882   | ug/l   | 97       |
| 22) Diisopropyl ether        | 3.763          | 45   | 152872     | 18.412   | ug/l   | 92       |
| 23) Vinyl Acetate            | 3.727          | 43   | 607653     | 84.369   | ug/l   | 99       |
| 24) 1,1-Dichloroethane       | 3.611          | 63   | 87491      | 19.830   | ug/l   | 98       |
| 25) 2-Butanone               | 4.562          | 43   | 192076     | 81.595   | ug/l   | 100      |
| 26) 2,2-Dichloropropane      | 4.471          | 77   | 63715      | 19.971   | ug/l   | 100      |
| 27) cis-1,2-Dichloroethene   | 4.489          | 96   | 54614      | 19.786   | ug/l   | 99       |
| 28) Bromochloromethane       | 4.897          | 49   | 39850      | 21.380   | ug/l   | 100      |
| 29) Tetrahydrofuran          | 5.013          | 42   | 122734     | 77.490   | ug/l   | 96       |
| 30) Chloroform               | 5.098          | 83   | 91933      | 21.045   | ug/l   | 95       |
| 31) Cyclohexane              | 5.476          | 56   | 69183      | 17.415   | ug/l   | 96       |
| 32) 1,1,1-Trichloroethane    | 5.385          | 97   | 76634      | 21.582   | ug/l   | 97       |
| 36) 1,1-Dichloropropene      | 5.696          | 75   | 61517      | 19.228   | ug/l   | 98       |
| 37) Ethyl Acetate            | 4.720          | 43   | 73212      | 16.159   | ug/l   | 99       |
| 38) Carbon Tetrachloride     | 5.678          | 117  | 64926      | 21.738   | ug/l   | 98       |
| 39) Methylcyclohexane        | 7.379          | 83   | 75515      | 18.286   | ug/l   | 93       |
| 40) Benzene                  | 6.043          | 78   | 194803     | 19.352   | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX082322\  
 Data File : VX030723.D  
 Acq On : 23 Aug 2022 11:44  
 Operator : JC/MD  
 Sample : VX0823WBS01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0823WBS01

Manual Integrations  
 APPROVED

Reviewed By :Krupa Patel 08/24/2022  
 Supervised By :Mahesh Dadoda 08/24/2022

Quant Time: Aug 24 00:45:43 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072522W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Jul 25 16:49:17 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 4.928  | 41   | 38945    | 16.036  | ug/l   | 95       |
| 42) 1,2-Dichloroethane        | 6.086  | 62   | 64863    | 18.648  | ug/l   | 99       |
| 43) Isopropyl Acetate         | 6.342  | 43   | 111320   | 16.761  | ug/l   | 98       |
| 44) Trichloroethene           | 7.129  | 130  | 54493    | 19.812  | ug/l   | 97       |
| 45) 1,2-Dichloropropane       | 7.433  | 63   | 52765    | 19.653  | ug/l   | 98       |
| 46) Dibromomethane            | 7.586  | 93   | 37598    | 20.593  | ug/l   | 94       |
| 47) Bromodichloromethane      | 7.824  | 83   | 74389    | 22.412  | ug/l   | 94       |
| 48) Methyl methacrylate       | 7.696  | 41   | 55526    | 16.985  | ug/l   | 95       |
| 49) 1,4-Dioxane               | 7.683  | 88   | 28866    | 382.306 | ug/l   | 97       |
| 51) 4-Methyl-2-Pentanone      | 8.573  | 43   | 375730   | 83.329  | ug/l   | 95       |
| 52) Toluene                   | 8.720  | 92   | 121040   | 19.709  | ug/l   | 98       |
| 53) t-1,3-Dichloropropene     | 8.982  | 75   | 72660    | 20.308  | ug/l   | 100      |
| 54) cis-1,3-Dichloropropene   | 8.366  | 75   | 81204    | 20.827  | ug/l   | 95       |
| 55) 1,1,2-Trichloroethane     | 9.153  | 97   | 54306    | 20.768  | ug/l   | 98       |
| 56) Ethyl methacrylate        | 9.122  | 69   | 79125    | 19.142  | ug/l   | 95       |
| 57) 1,3-Dichloropropane       | 9.311  | 76   | 86296    | 19.993  | ug/l   | 99       |
| 58) 2-Chloroethyl Vinyl ether | 8.244  | 63   | 215751   | 105.167 | ug/l   | 99       |
| 59) 2-Hexanone                | 9.433  | 43   | 288076   | 80.981  | ug/l   | 96       |
| 60) Dibromochloromethane      | 9.524  | 129  | 56567    | 23.369  | ug/l   | 100      |
| 61) 1,2-Dibromoethane         | 9.610  | 107  | 55088    | 20.176  | ug/l   | 98       |
| 64) Tetrachloroethene         | 9.275  | 164  | 48200    | 19.065  | ug/l   | 95       |
| 65) Chlorobenzene             | 10.079 | 112  | 133817   | 20.977  | ug/l   | 97       |
| 66) 1,1,1,2-Tetrachloroethane | 10.165 | 131  | 49170    | 22.674  | ug/l   | 99       |
| 67) Ethyl Benzene             | 10.195 | 91   | 234914   | 24.030  | ug/l   | 98       |
| 68) m/p-Xylenes               | 10.305 | 106  | 178339   | 41.268  | ug/l   | 96       |
| 69) o-Xylene                  | 10.646 | 106  | 87915    | 20.641  | ug/l   | 95       |
| 70) Styrene                   | 10.658 | 104  | 147951   | 21.143  | ug/l   | 99       |
| 71) Bromoform                 | 10.805 | 173  | 39278    | 24.637  | ug/l # | 96       |
| 73) Isopropylbenzene          | 10.963 | 105  | 230705   | 19.668  | ug/l   | 100      |
| 74) N-amyl acetate            | 10.841 | 43   | 89204    | 15.647  | ug/l   | 94       |
| 75) 1,1,2,2-Tetrachloroethane | 11.213 | 83   | 82064    | 19.114  | ug/l   | 100      |
| 76) 1,2,3-Trichloropropane    | 11.244 | 75   | 76685m   | 19.362  | ug/l   |          |
| 77) Bromobenzene              | 11.201 | 156  | 52621    | 19.113  | ug/l   | 93       |
| 78) n-propylbenzene           | 11.305 | 91   | 276059   | 19.719  | ug/l   | 99       |
| 79) 2-Chlorotoluene           | 11.366 | 91   | 169336   | 19.154  | ug/l   | 98       |
| 80) 1,3,5-Trimethylbenzene    | 11.451 | 105  | 198921   | 20.253  | ug/l   | 99       |
| 81) trans-1,4-Dichloro-2-b... | 11.018 | 75   | 26015    | 19.705  | ug/l   | 97       |
| 82) 4-Chlorotoluene           | 11.457 | 91   | 187522   | 19.413  | ug/l   | 100      |
| 83) tert-Butylbenzene         | 11.719 | 119  | 188528   | 20.709  | ug/l   | 98       |
| 84) 1,2,4-Trimethylbenzene    | 11.756 | 105  | 201917   | 20.606  | ug/l   | 99       |
| 85) sec-Butylbenzene          | 11.890 | 105  | 258732   | 20.764  | ug/l   | 97       |
| 86) p-Isopropyltoluene        | 12.012 | 119  | 211439   | 21.210  | ug/l   | 97       |
| 87) 1,3-Dichlorobenzene       | 11.969 | 146  | 104302   | 19.480  | ug/l   | 98       |
| 88) 1,4-Dichlorobenzene       | 12.042 | 146  | 110581   | 20.288  | ug/l   | 98       |
| 89) n-Butylbenzene            | 12.335 | 91   | 195966   | 21.114  | ug/l   | 99       |
| 90) Hexachloroethane          | 12.542 | 117  | 41904    | 24.414  | ug/l   | 93       |
| 91) 1,2-Dichlorobenzene       | 12.335 | 146  | 109591   | 20.853  | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.945 | 75   | 21096    | 20.487  | ug/l   | 90       |
| 93) 1,2,4-Trichlorobenzene    | 13.591 | 180  | 65809    | 21.593  | ug/l   | 95       |
| 94) Hexachlorobutadiene       | 13.725 | 225  | 26576    | 21.717  | ug/l   | 99       |
| 95) Naphthalene               | 13.780 | 128  | 250861   | 21.072  | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 13.963 | 180  | 68582    | 21.942  | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX082322\  
Data File : VX030723.D  
Acq On : 23 Aug 2022 11:44  
Operator : JC/MD  
Sample : VX0823WBS01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01

Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022

Quant Time: Aug 24 00:45:43 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072522W.M  
Quant Title : SW846 8260  
QLast Update : Mon Jul 25 16:49:17 2022  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|


-----  
(#) = qualifier out of range (m) = manual integration (+) = signals summed





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.556 min Scan# 71  
Delta R.T. -0.006 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion:168 Resp: 13613  
Ion Ratio Lower Upper  
168 100  
99 57.7 44.0 66.0




Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#2  
Dichlorodifluoromethane  
Concen: 20.794 ug/l  
RT: 1.166 min Scan# 13  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44


Tgt Ion: 85 Resp: 46450  
Ion Ratio Lower Upper  
85 100  
87 31.3 16.5 49.5





#3  
Chloromethane  
Concen: 19.238 ug/l  
RT: 1.288 min Scan# 31  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion: 50 Resp: 42490  
Ion Ratio Lower Upper  
50 100  
52 34.1 26.3 39.5




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#4  
Vinyl Chloride  
Concen: 20.382 ug/l  
RT: 1.374 min Scan# 47  
Delta R.T. 0.001 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44

Tgt Ion: 62 Resp: 45441  
Ion Ratio Lower Upper  
62 100  
64 33.5 25.2 37.8









#5  
Bromomethane  
Concen: 15.834 ug/l  
RT: 1.611 min Scan# 80  
Delta R.T. 0.012 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01

Tgt Ion: 94 Resp: 1545  
Ion Ratio Lower Upper  
94 100  
96 93.6 76.9 115.3




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#6  
Chloroethane  
Concen: 22.010 ug/l  
RT: 1.684 min Scan# 98  
Delta R.T. 0.006 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44

Tgt Ion: 64 Resp: 28176  
Ion Ratio Lower Upper  
64 100  
66 33.4 27.2 40.8









#7  
 Trichlorofluoromethane  
 Concen: 22.760 ug/l  
 RT: 1.886 min Scan# 11  
 Delta R.T. 0.007 min  
 Lab File: VX030723.D  
 Acq: 23 Aug 2022 11:44

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0823WBS01

Tgt Ion:101 Resp: 66854  
 Ion Ratio Lower Upper  
 101 100  
 103 65.6 49.7 74.5




Manual Integrations  
 APPROVED


Reviewed By :Krupa Patel 08/24/2022  
 Supervised By :Mahesh Dadoda 08/24/2022



#8  
 Diethyl Ether  
 Concen: 23.668 ug/l  
 RT: 2.136 min Scan# 172  
 Delta R.T. 0.001 min  
 Lab File: VX030723.D  
 Acq: 23 Aug 2022 11:44


Tgt Ion: 74 Resp: 27224  
 Ion Ratio Lower Upper  
 74 100  
 45 97.7 51.5 154.7





#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 20.847 ug/l  
RT: 2.325 min Scan# 203  
Delta R.T. 0.001 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01




Tgt Ion:101 Resp: 4535  
Ion Ratio Lower Upper  
101 100  
85 42.0 34.0 51.0  
151 73.4 59.4 89.0

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#10  
Methyl Iodide  
Concen: 13.532 ug/l  
RT: 2.453 min Scan# 224  
Delta R.T. 0.007 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44



Tgt Ion:142 Resp: 26260  
Ion Ratio Lower Upper  
142 100  
127 42.6 30.2 45.2  
141 13.4 11.3 16.9









#11  
Tert butyl alcohol  
Concen: 97.436 ug/l  
RT: 2.983 min Scan# 310  
Delta R.T. 0.006 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01

Tgt Ion: 59 Resp: 6503  
Ion Ratio Lower Upper  
59 100  
57 8.8 7.4 11.2




Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#12  
1,1-Dichloroethene  
Concen: 19.065 ug/l  
RT: 2.318 min Scan# 202  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44


Tgt Ion: 96 Resp: 40543  
Ion Ratio Lower Upper  
96 100  
61 146.6 117.4 176.0  
98 73.6 52.2 78.4





#13  
Acrolein  
Concen: 19.019 ug/l  
RT: 2.239 min Scan# 189  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion: 56 Resp: 3120  
Ion Ratio Lower Upper  
56 100  
55 75.6 56.6 84.8




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#14  
Allyl chloride  
Concen: 16.534 ug/l  
RT: 2.666 min Scan# 259  
Delta R.T. -0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44

Tgt Ion: 41 Resp: 72556  
Ion Ratio Lower Upper  
41 100  
39 57.6 48.9 73.3  
76 35.7 26.9 40.3









#15  
Acrylonitrile  
Concen: 88.305 ug/l  
RT: 3.068 min Scan# 31  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01

Tgt Ion: 53 Resp: 14564  
Ion Ratio Lower Upper  
53 100  
52 83.6 65.4 98.0  
51 35.5 27.7 41.5




Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#16  
Acetone  
Concen: 79.924 ug/l  
RT: 2.386 min Scan# 213  
Delta R.T. 0.001 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44


Tgt Ion: 43 Resp: 111961  
Ion Ratio Lower Upper  
43 100  
58 32.6 27.7 41.5





#17  
Carbon Disulfide  
Concen: 17.280 ug/l  
RT: 2.507 min Scan# 211  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion: 76 Resp: 9919  
Ion Ratio Lower Upper  
76 100  
78 9.1 7.8 11.6




Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#18  
Methyl Acetate  
Concen: 17.762 ug/l  
RT: 2.709 min Scan# 266  
Delta R.T. 0.001 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44


Tgt Ion: 43 Resp: 73594  
Ion Ratio Lower Upper  
43 100  
74 25.0 19.4 29.2





#19  
Methyl tert-butyl Ether  
Concen: 20.048 ug/l  
RT: 3.117 min Scan# 311  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
Client Sample Id :  
VX0823WBS01



Tgt Ion: 73 Resp: 14604  
Ion Ratio Lower Upper  
73 100  
57 22.2 19.0 28.6




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#20  
Methylene Chloride  
Concen: 20.073 ug/l  
RT: 2.788 min Scan# 279  
Delta R.T. -0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44


Tgt Ion: 84 Resp: 51842  
Ion Ratio Lower Upper  
84 100  
49 112.4 93.8 140.6  
51 32.2 29.4 44.0  
86 60.9 51.0 76.4





#21  
trans-1,2-Dichloroethene  
Concen: 18.882 ug/l  
RT: 3.093 min Scan# 31  
Delta R.T. 0.001 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion: 96 Resp: 4432  
Ion Ratio Lower Upper  
96 100  
61 130.2 105.6 158.4  
98 66.4 50.1 75.1




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#22  
Diisopropyl ether  
Concen: 18.412 ug/l  
RT: 3.763 min Scan# 439  
Delta R.T. -0.006 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44


Tgt Ion: 45 Resp: 152872  
Ion Ratio Lower Upper  
45 100  
43 73.0 65.4 98.0  
87 29.6 22.3 33.5  
59 12.1 9.0 13.4





#23  
 Vinyl Acetate  
 Concen: 84.369 ug/l  
 RT: 3.727 min Scan# 41  
 Delta R.T. -0.000 min  
 Lab File: VX030723.D  
 Acq: 23 Aug 2022 11:44




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0823WBS01



Tgt Ion: 43 Resp: 60765  
 Ion Ratio Lower Upper  
 43 100  
 86 11.1 8.5 12.7




Manual Integrations  
 APPROVED


Reviewed By :Krupa Patel 08/24/2022  
 Supervised By :Mahesh Dadoda 08/24/2022



#24  
 1,1-Dichloroethane  
 Concen: 19.830 ug/l  
 RT: 3.611 min Scan# 414  
 Delta R.T. -0.000 min  
 Lab File: VX030723.D  
 Acq: 23 Aug 2022 11:44


Tgt Ion: 63 Resp: 87491  
 Ion Ratio Lower Upper  
 63 100  
 98 6.3 3.5 10.5  
 100 4.9 2.3 6.9





#25  
2-Butanone  
Concen: 81.595 ug/l  
RT: 4.562 min Scan# 51  
Delta R.T. -0.006 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion: 43 Resp: 192070  
Ion Ratio Lower Upper  
43 100  
72 27.0 21.7 32.5




Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#26  
2,2-Dichloropropane  
Concen: 19.971 ug/l  
RT: 4.471 min Scan# 555  
Delta R.T. -0.005 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44

Tgt Ion: 77 Resp: 63715  
Ion Ratio Lower Upper  
77 100  
97 26.0 13.0 39.0





#27

cis-1,2-Dichloroethene

Concen: 19.786 ug/l

RT: 4.489 min Scan# 51

Delta R.T. -0.006 min

Lab File: VX030723.D

Acq: 23 Aug 2022 11:44

Instrument :

MSVOA\_X

ClientSampleId :

VX0823WBS01

Tgt Ion: 96 Resp: 54614





| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 96  | 100   |       |       |
| 61  | 134.4 | 0.0   | 270.0 |
| 98  | 65.8  | 0.0   | 127.8 |

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/24/2022

Supervised By :Mahesh Dadoda 08/24/2022



#28

Bromochloromethane

Concen: 21.380 ug/l

RT: 4.897 min Scan# 625




Delta R.T. -0.006 min

Lab File: VX030723.D

Acq: 23 Aug 2022 11:44

Tgt Ion: 49 Resp: 39850

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 49  | 100   |       |       |
| 129 | 2.1   | 0.0   | 4.4   |
| 130 | 78.5  | 62.8  | 94.2  |





#29

Tetrahydrofuran

Concen: 77.490 ug/l

RT: 5.013 min Scan# 644

Delta R.T. 0.000 min

Lab File: VX030723.D


Acq: 23 Aug 2022 11:44

Instrument :

MSVOA\_X

ClientSampleId :

VX0823WBS01






| Tgt Ion:  | 42   | Resp: | 122734 |
|-----------|------|-------|--------|
| Ion Ratio | 100  | Lower | Upper  |
| 42        | 100  |       |        |
| 72        | 48.5 | 36.5  | 54.7   |
| 71        | 44.7 | 33.9  | 50.9   |

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/24/2022

Supervised By :Mahesh Dadoda 08/24/2022



#30

Chloroform

Concen: 21.045 ug/l




RT: 5.098 min Scan# 658


Delta R.T. 0.000 min

Lab File: VX030723.D

Acq: 23 Aug 2022 11:44


| Tgt Ion:  | 83   | Resp: | 91933 |
|-----------|------|-------|-------|
| Ion Ratio | 100  | Lower | Upper |
| 83        | 100  |       |       |
| 85        | 60.5 | 51.5  | 77.3  |





#31  
Cyclohexane  
Concen: 17.415 ug/l  
RT: 5.476 min Scan# 719  
Delta R.T. 0.006 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01




Tgt Ion: 56 Resp: 6918  
Ion Ratio Lower Upper  
56 100  
69 28.5 25.4 38.0  
84 83.6 68.9 103.3

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022




#32  
1,1,1-Trichloroethane  
Concen: 21.582 ug/l  
RT: 5.385 min Scan# 705  
Delta R.T. -0.006 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44



Tgt Ion: 97 Resp: 76634  
Ion Ratio Lower Upper  
97 100  
99 63.1 51.4 77.0  
61 42.0 35.8 53.8









#33  
1,2-Dichloroethane-d4  
Concen: 48.122 ug/l  
RT: 5.958 min Scan# 799  
Delta R.T. -0.006 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01

Tgt Ion: 65 Resp: 131669  
Ion Ratio Lower Upper  
65 100  
67 55.2 0.0 109.6




Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 6.769 min Scan# 932  
Delta R.T. -0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44


Tgt Ion:114 Resp: 232560  
Ion Ratio Lower Upper  
114 100  
63 17.9 0.0 38.4  
88 15.0 0.0 31.8





#35  
Dibromofluoromethane  
Concen: 53.547 ug/l  
RT: 5.391 min Scan# 706  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion: 113 Resp: 119044  
Ion Ratio Lower Upper  
113 100  
111 102.6 83.2 124.8  
192 16.9 14.8 22.2




Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#36  
1,1-Dichloropropene  
Concen: 19.228 ug/l  
RT: 5.696 min Scan# 756  
Delta R.T. -0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44


Tgt Ion: 75 Resp: 61517  
Ion Ratio Lower Upper  
75 100  
110 37.2 19.3 57.9  
77 31.6 24.4 36.6





#37  
Ethyl Acetate  
Concen: 16.159 ug/l  
RT: 4.720 min Scan# 596  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01




Tgt Ion: 43 Resp: 73212  
Ion Ratio Lower Upper  
43 100  
61 14.9 11.7 17.5  
70 11.4 9.3 13.9

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022




#38  
Carbon Tetrachloride  
Concen: 21.738 ug/l  
RT: 5.678 min Scan# 753  
Delta R.T. -0.005 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Tgt Ion:117 Resp: 64926  
Ion Ratio Lower Upper  
117 100  
119 91.4 74.1 111.1  
121 30.1 25.8 38.8





#39  
Methylcyclohexane  
Concen: 18.286 ug/l  
RT: 7.379 min Scan# 1032  
Delta R.T. -0.005 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion: 83 Resp: 7551  
Ion Ratio Lower Upper  
83 100  
55 81.1 57.7 86.5  
98 47.0 36.5 54.7




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#40  
Benzene  
Concen: 19.352 ug/l  
RT: 6.043 min Scan# 813  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44


Tgt Ion: 78 Resp: 194803  
Ion Ratio Lower Upper  
78 100  
77 23.1 18.0 27.0





#41  
Methacrylonitrile  
Concen: 16.036 ug/l  
RT: 4.928 min Scan# 61  
Delta R.T. -0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion: 41 Resp: 3894  
Ion Ratio Lower Upper  
41 100  
39 55.6 42.9 64.3  
67 81.0 60.1 90.1  
52 31.5 22.6 34.0




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#42  
1,2-Dichloroethane  
Concen: 18.648 ug/l  
RT: 6.086 min Scan# 820  
Delta R.T. -0.006 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44


Tgt Ion: 62 Resp: 64863  
Ion Ratio Lower Upper  
62 100  
98 10.3 0.0 20.2





#43  
Isopropyl Acetate  
Concen: 16.761 ug/l  
RT: 6.342 min Scan# 862  
Delta R.T. -0.006 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion: 43 Resp: 111320  
Ion Ratio Lower Upper  
43 100  
61 21.5 16.5 24.7  
87 14.5 11.1 16.7




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#44  
Trichloroethene  
Concen: 19.812 ug/l  
RT: 7.129 min Scan# 991  
Delta R.T. 0.001 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44

Tgt Ion:130 Resp: 54493  
Ion Ratio Lower Upper  
130 100  
95 100.1 0.0 194.8





#45

1,2-Dichloropropane

Concen: 19.653 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. 0.000 min

Lab File: VX030723.D

Acq: 23 Aug 2022 11:44

Instrument :

MSVOA\_X

ClientSampleId :

VX0823WBS01

Tgt Ion: 63 Resp: 5276

Ion Ratio Lower Upper

63 100





65 32.4 25.2 37.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/24/2022

Supervised By :Mahesh Dadoda 08/24/2022



#46

Dibromomethane

Concen: 20.593 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX030723.D

Acq: 23 Aug 2022 11:44




Tgt Ion: 93 Resp: 37598


Ion Ratio Lower Upper

93 100

95 81.6 67.4 101.0

174 86.2 75.6 113.4









#47  
Bromodichloromethane  
Concen: 22.412 ug/l  
RT: 7.824 min Scan# 1105  
Delta R.T. -0.005 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01

Tgt Ion: 83 Resp: 74389  
Ion Ratio Lower Upper  
83 100  
85 61.1 53.5 80.3  
127 8.4 6.6 10.0




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#48  
Methyl methacrylate  
Concen: 16.985 ug/l  
RT: 7.696 min Scan# 1084  
Delta R.T. 0.001 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44

Tgt Ion: 41 Resp: 55526  
Ion Ratio Lower Upper  
41 100  
69 90.0 68.8 103.2  
39 50.9 44.6 66.8





#49

1,4-Dioxane

Concen: 382.306 ug/l

RT: 7.683 min Scan# 1082

Delta R.T. 0.018 min

Lab File: VX030723.D

Acq: 23 Aug 2022 11:44

Instrument :

MSVOA\_X

ClientSampleId :

VX0823WBS01





|             |             |
|-------------|-------------|
| Tgt Ion: 88 | Resp: 2886  |
| Ion Ratio   | Lower Upper |
| 88 100      |             |
| 43 28.2     | 23.8 35.8   |
| 58 69.1     | 53.8 80.6   |

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/24/2022

Supervised By :Mahesh Dadoda 08/24/2022



#50

Toluene-d8

Concen: 50.395 ug/l




RT: 8.653 min Scan# 1241


Delta R.T. 0.001 min

Lab File: VX030723.D

Acq: 23 Aug 2022 11:44

|             |              |
|-------------|--------------|
| Tgt Ion: 98 | Resp: 433872 |
| Ion Ratio   | Lower Upper  |
| 98 100      |              |
| 100 64.9    | 52.3 78.5    |





#51

4-Methyl-2-Pentanone

Concen: 83.329 ug/l

RT: 8.573 min Scan# 1228

Delta R.T. -0.006 min

Lab File: VX030723.D

Acq: 23 Aug 2022 11:44

Instrument :

MSVOA\_X

ClientSampleId :

VX0823WBS01

Tgt Ion: 43 Resp: 375730

Ion Ratio Lower Upper

43 100





58 42.3 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/24/2022

Supervised By :Mahesh Dadoda 08/24/2022



#52

Toluene

Concen: 19.709 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. -0.000 min

Lab File: VX030723.D




Acq: 23 Aug 2022 11:44


Tgt Ion: 92 Resp: 121040

Ion Ratio Lower Upper

92 100


91 176.5 139.2 208.8





#53  
t-1,3-Dichloropropene  
Concen: 20.308 ug/l  
RT: 8.982 min Scan# 1194  
Delta R.T. -0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion: 75 Resp: 72660  
Ion Ratio Lower Upper  
75 100  
77 32.7 26.3 39.5




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#54  
cis-1,3-Dichloropropene  
Concen: 20.827 ug/l  
RT: 8.366 min Scan# 1194  
Delta R.T. -0.006 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44

Tgt Ion: 75 Resp: 81204  
Ion Ratio Lower Upper  
75 100  
77 32.8 25.0 37.4  
39 42.5 37.5 56.3





#55

1,1,2-Trichloroethane

Concen: 20.768 ug/l

RT: 9.153 min Scan# 1324

Delta R.T. -0.005 min

Lab File: VX030723.D

Acq: 23 Aug 2022 11:44

Instrument :

MSVOA\_X

ClientSampleId :

VX0823WBS01

Tgt Ion: 97 Resp: 54300

Ion Ratio Lower Upper

97 100

83 82.2 68.9 103.3

85 54.7 44.0 66.0





99 61.2 48.3 72.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/24/2022

Supervised By :Mahesh Dadoda 08/24/2022



#56

Ethyl methacrylate

Concen: 19.142 ug/l

RT: 9.122 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VX030723.D

Acq: 23 Aug 2022 11:44




Tgt Ion: 69 Resp: 79125


Ion Ratio Lower Upper

69 100

41 62.8 52.3 78.5

39 30.0 28.2 42.4





#57

1,3-Dichloropropane

Concen: 19.993 ug/l

RT: 9.311 min Scan# 1149

Delta R.T. 0.000 min

Lab File: VX030723.D

Acq: 23 Aug 2022 11:44

Instrument :

MSVOA\_X

ClientSampleId :

VX0823WBS01

Tgt Ion: 76 Resp: 86290

Ion Ratio Lower Upper

76 100





78 33.1 25.9 38.9

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/24/2022

Supervised By :Mahesh Dadoda 08/24/2022



#58

2-Chloroethyl Vinyl ether

Concen: 105.167 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.006 min

Lab File: VX030723.D




Acq: 23 Aug 2022 11:44


Tgt Ion: 63 Resp: 215751

Ion Ratio Lower Upper

63 100


106 28.0 22.2 33.2





#59  
2-Hexanone  
Concen: 80.981 ug/l  
RT: 9.433 min Scan# 1369  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion: 43 Resp: 288070  
Ion Ratio Lower Upper  
43 100  
58 57.8 27.6 82.8




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#60  
Dibromochloromethane  
Concen: 23.369 ug/l  
RT: 9.524 min Scan# 1384  
Delta R.T. 0.001 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44

Tgt Ion:129 Resp: 56567  
Ion Ratio Lower Upper  
129 100  
127 77.1 38.5 115.5









#61  
1,2-Dibromoethane  
Concen: 20.176 ug/l  
RT: 9.610 min Scan# 1399  
Delta R.T. -0.006 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01

Tgt Ion:107 Resp: 55088  
Ion Ratio Lower Upper  
107 100  
109 94.2 76.6 114.8




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#62  
4-Bromofluorobenzene  
Concen: 49.997 ug/l  
RT: 11.085 min Scan# 1640  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44


Tgt Ion: 95 Resp: 159751  
Ion Ratio Lower Upper  
95 100  
174 74.0 0.0 149.0  
176 70.7 0.0 146.0





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. -0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion:117 Resp: 234839  
Ion Ratio Lower Upper  
117 100  
82 55.3 42.7 64.1  
119 31.6 25.4 38.2




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#64  
Tetrachloroethene  
Concen: 19.065 ug/l  
RT: 9.275 min Scan# 1343  
Delta R.T. 0.001 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44


Tgt Ion:164 Resp: 48200  
Ion Ratio Lower Upper  
164 100  
166 127.6 98.2 147.4  
129 103.1 80.8 121.2  
131 104.4 75.5 113.3





#65  
Chlorobenzene  
Concen: 20.977 ug/l  
RT: 10.079 min Scan# 1475  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01




Tgt Ion:112 Resp: 13381  
Ion Ratio Lower Upper  
112 100  
114 32.1 24.6 36.8

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022




#66  
1,1,1,2-Tetrachloroethane  
Concen: 22.674 ug/l  
RT: 10.165 min Scan# 1489  
Delta R.T. 0.001 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Tgt Ion:131 Resp: 49170  
Ion Ratio Lower Upper  
131 100  
133 94.9 46.9 140.8  
119 69.9 34.4 103.1





#67  
Ethyl Benzene  
Concen: 24.030 ug/l  
RT: 10.195 min Scan# 1494  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01




Tgt Ion: 91 Resp: 234914  
Ion Ratio Lower Upper  
91 100  
106 30.5 25.5 38.3

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022




#68  
m/p-Xylenes  
Concen: 41.268 ug/l  
RT: 10.305 min Scan# 1512  
Delta R.T. -0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Tgt Ion:106 Resp: 178339  
Ion Ratio Lower Upper  
106 100  
91 203.8 158.2 237.4





#69  
o-Xylene  
Concen: 20.641 ug/l  
RT: 10.646 min Scan# 1568  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion:106 Resp: 87915  
Ion Ratio Lower Upper  
106 100  
91 215.1 103.8 311.6




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#70  
Styrene  
Concen: 21.143 ug/l  
RT: 10.658 min Scan# 1570  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44


Tgt Ion:104 Resp: 147951  
Ion Ratio Lower Upper  
104 100  
78 48.2 37.8 56.8  
103 55.0 44.5 66.7





#71  
Bromoform  
Concen: 24.637 ug/l  
RT: 10.805 min Scan# 1594  
Delta R.T. -0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion:173 Resp: 39278  
Ion Ratio Lower Upper  
173 100  
175 50.4 23.8 71.2  
254 0.0 0.1 0.1




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.024 min Scan# 1794  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44


Tgt Ion:152 Resp: 150255  
Ion Ratio Lower Upper  
152 100  
115 72.3 40.5 121.5  
150 165.0 0.0 347.0





#73  
Isopropylbenzene  
Concen: 19.668 ug/l  
RT: 10.963 min Scan# 1620  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01




Tgt Ion:105 Resp: 23070  
Ion Ratio Lower Upper  
105 100  
120 27.1 13.5 40.4

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022




#74  
N-ethyl acetate  
Concen: 15.647 ug/l  
RT: 10.841 min Scan# 1600  
Delta R.T. -0.006 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44



Tgt Ion: 43 Resp: 89204  
Ion Ratio Lower Upper  
43 100  
70 45.2 32.3 48.5  
55 26.7 18.4 27.6  
61 25.2 19.2 28.8









#75  
1,1,2,2-Tetrachloroethane  
Concen: 19.114 ug/l  
RT: 11.213 min Scan# 1661  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01

Tgt Ion: 83 Resp: 82064  
Ion Ratio Lower Upper  
83 100  
131 9.4 4.5 13.7  
85 64.8 32.5 97.5




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#76  
1,2,3-Trichloropropane  
Concen: 19.362 ug/l m  
RT: 11.244 min Scan# 1666  
Delta R.T. 0.001 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44


Tgt Ion: 75 Resp: 76685  
Ion Ratio Lower Upper  
75 100  
77 38.3 19.8 59.3





#77  
Bromobenzene  
Concen: 19.113 ug/l  
RT: 11.201 min Scan# 1659  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01




Tgt Ion:156 Resp: 5262  
Ion Ratio Lower Upper  
156 100  
77 176.2 82.7 247.9  
158 100.6 47.5 142.5

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022




#78  
n-propylbenzene  
Concen: 19.719 ug/l  
RT: 11.305 min Scan# 1676  
Delta R.T. -0.005 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Tgt Ion: 91 Resp: 276059  
Ion Ratio Lower Upper  
91 100  
120 22.5 11.6 34.8





#79  
2-Chlorotoluene  
Concen: 19.154 ug/l  
RT: 11.366 min Scan# 1686  
Delta R.T. 0.001 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01




Tgt Ion: 91 Resp: 169336  
Ion Ratio Lower Upper  
91 100  
126 33.8 17.4 52.3

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022




#80  
1,3,5-Trimethylbenzene  
Concen: 20.253 ug/l  
RT: 11.451 min Scan# 1700  
Delta R.T. -0.006 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Tgt Ion:105 Resp: 198921  
Ion Ratio Lower Upper  
105 100  
120 48.4 24.4 73.2





#81  
trans-1,4-Dichloro-2-butene  
Concen: 19.705 ug/l  
RT: 11.018 min Scan# 1630  
Delta R.T. -0.006 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion: 75 Resp: 26015  
Ion Ratio Lower Upper  
75 100  
53 88.2 74.2 111.4  
89 43.9 34.8 52.2




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#82  
4-Chlorotoluene  
Concen: 19.413 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44


Tgt Ion: 91 Resp: 187522  
Ion Ratio Lower Upper  
91 100  
126 30.6 15.3 45.9





#83  
tert-Butylbenzene  
Concen: 20.709 ug/l  
RT: 11.719 min Scan# 1744  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion:119 Resp: 188528  
Ion Ratio Lower Upper  
119 100  
91 58.6 28.3 84.9  
134 22.6 11.2 33.5




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#84  
1,2,4-Trimethylbenzene  
Concen: 20.606 ug/l  
RT: 11.756 min Scan# 1750  
Delta R.T. -0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44


Tgt Ion:105 Resp: 201917  
Ion Ratio Lower Upper  
105 100  
120 45.0 22.9 68.8





#85  
 sec-Butylbenzene  
 Concen: 20.764 ug/l  
 RT: 11.890 min Scan# 1772  
 Delta R.T. -0.006 min  
 Lab File: VX030723.D  
 Acq: 23 Aug 2022 11:44




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0823WBS01




Tgt Ion:105 Resp: 25873  
 Ion Ratio Lower Upper  
 105 100  
 134 18.7 10.0 30.0

Manual Integrations  
 APPROVED



Reviewed By :Krupa Patel 08/24/2022  
 Supervised By :Mahesh Dadoda 08/24/2022




#86  
 p-Isopropyltoluene  
 Concen: 21.210 ug/l  
 RT: 12.012 min Scan# 1792  
 Delta R.T. -0.000 min  
 Lab File: VX030723.D  
 Acq: 23 Aug 2022 11:44



Tgt Ion:119 Resp: 211439  
 Ion Ratio Lower Upper  
 119 100  
 134 25.5 13.9 41.6  
 91 23.9 11.8 35.3





#87

1,3-Dichlorobenzene

Concen: 19.480 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.006 min

Lab File: VX030723.D

Acq: 23 Aug 2022 11:44

Instrument :

MSVOA\_X

ClientSampleId :

VX0823WBS01

Tgt Ion:146 Resp: 104302

Ion Ratio Lower Upper

146 100

111 41.5 20.3 60.8





148 65.0 31.8 95.3

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/24/2022

Supervised By :Mahesh Dadoda 08/24/2022



#88

1,4-Dichlorobenzene

Concen: 20.288 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. 0.000 min

Lab File: VX030723.D

Acq: 23 Aug 2022 11:44




Tgt Ion:146 Resp: 110581

Ion Ratio Lower Upper

146 100

111 42.2 20.3 60.9


148 63.0 32.3 96.8





#89  
n-Butylbenzene  
Concen: 21.114 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. -0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion: 91 Resp: 195960  
Ion Ratio Lower Upper  
91 100  
92 52.5 26.2 78.5  
134 23.7 12.6 37.8




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#90  
Hexachloroethane  
Concen: 24.414 ug/l  
RT: 12.542 min Scan# 1879  
Delta R.T. 0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44


Tgt Ion:117 Resp: 41904  
Ion Ratio Lower Upper  
117 100  
201 62.8 34.2 102.5





#91  
1,2-Dichlorobenzene  
Concen: 20.853 ug/l  
RT: 12.335 min Scan# 1846  
Delta R.T. -0.006 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion:146 Resp: 10959  
Ion Ratio Lower Upper  
146 100  
111 43.5 20.7 62.1  
148 63.2 32.0 96.2




Manual Integrations  
APPROVED


Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#92  
1,2-Dibromo-3-Chloropropane  
Concen: 20.487 ug/l  
RT: 12.945 min Scan# 1945  
Delta R.T. 0.001 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44


Tgt Ion: 75 Resp: 21096  
Ion Ratio Lower Upper  
75 100  
155 78.6 45.0 134.8  
157 104.4 56.9 170.7





#93  
1,2,4-Trichlorobenzene  
Concen: 21.593 ug/l  
RT: 13.591 min Scan# 2051  
Delta R.T. -0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01




Tgt Ion:180 Resp: 65809  
Ion Ratio Lower Upper  
180 100  
182 90.7 48.2 144.6  
145 32.6 15.9 47.6

Manual Integrations  
APPROVED



Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022




#94  
Hexachlorobutadiene  
Concen: 21.717 ug/l  
RT: 13.725 min Scan# 2073  
Delta R.T. -0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Tgt Ion:225 Resp: 26576  
Ion Ratio Lower Upper  
225 100  
223 62.8 30.6 91.6  
227 63.9 31.9 95.9





#95  
Naphthalene  
Concen: 21.072 ug/l  
RT: 13.780 min Scan# 2082  
Delta R.T. -0.000 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0823WBS01



Tgt Ion:128 Resp: 25086  
Ion Ratio Lower Upper  
128 100  
127 13.2 10.2 15.2  
129 11.2 8.6 13.0

Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 08/24/2022  
Supervised By :Mahesh Dadoda 08/24/2022



#96  
1,2,3-Trichlorobenzene  
Concen: 21.942 ug/l  
RT: 13.963 min Scan# 2112  
Delta R.T. 0.001 min  
Lab File: VX030723.D  
Acq: 23 Aug 2022 11:44

Tgt Ion:180 Resp: 68582  
Ion Ratio Lower Upper  
180 100  
182 95.1 47.0 141.0  
145 35.2 17.0 50.9

