

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083119\
 Data File : VX011949.D
 Acq On : 31 Aug 2019 14:27
 Operator : SY/SP
 Sample : K4516-13
 Misc : 6.69g/5.0mL/MSVOA X/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T2-S-B3-(0)

Quant Time: Sep 01 02:33:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082919S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 01:59:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	433195	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	576786	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	495453	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	242171	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	181191	43.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.32%	
35) Dibromofluoromethane	5.48	113	166901	45.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.26%	
50) Toluene-d8	8.71	98	652193	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
62) 4-Bromofluorobenzene	11.14	95	226464	39.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.10%	
Target Compounds						
16) Acetone	2.43	43	30714	34.394	ug/l	99
29) Tetrahydrofuran	5.12	42	4349	5.357	ug/l #	81
30) Chloroform	5.19	83	10660	1.552	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083119\
Data File : VX011949.D
Acq On : 31 Aug 2019 14:27
Operator : SY/SP
Sample : K4516-13
Misc : 6.69g/5.0mL/MSVOA X/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T2-S-B3-(0)

Quant Time: Sep 01 02:33:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082919S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 01:59:05 2019
Response via : Initial Calibration

