

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083119\
 Data File : VX011950.D
 Acq On : 31 Aug 2019 14:55
 Operator : SY/SP
 Sample : K4516-14
 Misc : 6.24g/5.0mL/MSVOA X/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T2-S-B3-(2.5)

Quant Time: Sep 01 02:36:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082919S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 01:59:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	410541	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	545659	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	471831	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	239510	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	170390	43.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.64%	
35) Dibromofluoromethane	5.48	113	160145	45.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.54%	
50) Toluene-d8	8.71	98	611731	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
62) 4-Bromofluorobenzene	11.13	95	221511	40.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.74%	
Target Compounds						
16) Acetone	2.43	43	48390	57.178	ug/l	99
29) Tetrahydrofuran	5.14	42	4216	5.480	ug/l	94
30) Chloroform	5.20	83	13164	2.022	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083119\
Data File : VX011950.D
Acq On : 31 Aug 2019 14:55
Operator : SY/SP
Sample : K4516-14
Misc : 6.24g/5.0mL/MSVOA X/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T2-S-B3-(2.5)

Quant Time: Sep 01 02:36:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082919S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 01:59:05 2019
Response via : Initial Calibration

