

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083119\
 Data File : VX011969.D
 Acq On : 01 Sep 2019 00:44
 Operator : SY/SP
 Sample : K4526-11
 Misc : 5.45g/5.0mL/MSVOA X/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T9-12-13-F2-(1.9)

Quant Time: Sep 01 03:59:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082919S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 01:59:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	431304	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	544444	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	478413	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	254839	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	169464	41.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.02%	
35) Dibromofluoromethane	5.48	113	157228	45.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.08%	
50) Toluene-d8	8.71	98	609496	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
62) 4-Bromofluorobenzene	11.13	95	228359	41.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.42%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.02	59	3578	9.000	ug/l	96
16) Acetone	2.43	43	62568	70.372	ug/l	99
20) Methylene Chloride	2.84	84	37831	5.613	ug/l	95
29) Tetrahydrofuran	5.12	42	4718	5.837	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083119\
Data File : VX011969.D
Acq On : 01 Sep 2019 00:44
Operator : SY/SP
Sample : K4526-11
Misc : 5.45g/5.0mL/MSVOA X/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T9-12-13-F2-(1.9)

Quant Time: Sep 01 03:59:06 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082919S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 01:59:05 2019
Response via : Initial Calibration

