

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083119\
 Data File : VX011971.D
 Acq On : 01 Sep 2019 01:41
 Operator : SY/SP
 Sample : K4526-13
 Misc : 3.03g/5.0mL/MSVOA X/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T9-12-13-F4-(1.9)

Quant Time: Sep 01 04:11:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082919S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 01:59:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	429135	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	545470	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	468048	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	230421	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	168967	41.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.20%	
35) Dibromofluoromethane	5.48	113	156394	44.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.42%	
50) Toluene-d8	8.71	98	604388	48.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.68%	
62) 4-Bromofluorobenzene	11.14	95	214954	39.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.38%	
Target Compounds						
29) Tetrahydrofuran	5.12	42	4502	5.598	ug/l #	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083119\
Data File : VX011971.D
Acq On : 01 Sep 2019 01:41
Operator : SY/SP
Sample : K4526-13
Misc : 3.03g/5.0mL/MSVOA X/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T9-12-13-F4-(1.9)

Quant Time: Sep 01 04:11:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082919S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 01:59:05 2019
Response via : Initial Calibration

