

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX090222\
 Data File : VX031078.D
 Acq On : 02 Sep 2022 19:32
 Operator : JC/MD
 Sample : N4297-18MSD
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-MW01D2-20220821MSD

Quant Time: Sep 03 00:51:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	174156	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	278172	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	256472	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	131776	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	109772	48.987	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.980%
35) Dibromofluoromethane	5.391	113	105310	51.877	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.760%
50) Toluene-d8	8.653	98	372675	49.882	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.760%
62) 4-Bromofluorobenzene	11.085	95	138646	52.542	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.080%

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	1.166	85	75602	42.189	ug/l	99
3) Chloromethane	1.288	50	59142	39.204	ug/l	100
4) Vinyl Chloride	1.374	62	78343	41.852	ug/l	99
5) Bromomethane	1.611	94	44954	53.717	ug/l	95
6) Chloroethane	1.684	64	31966	28.307	ug/l	99
7) Trichlorofluoromethane	1.886	101	153695	48.590	ug/l	98
8) Diethyl Ether	2.136	74	49778	45.123	ug/l	70
9) 1,1,2-Trichlorotrifluo...	2.331	101	93384	44.572	ug/l	93
10) Methyl Iodide	2.453	142	111172	54.546	ug/l	95
11) Tert butyl alcohol	3.007	59	123684	275.346	ug/l	99
12) 1,1-Dichloroethene	2.319	96	96057	47.477	ug/l	92
13) Acrolein	2.239	56	100715	245.611	ug/l	99
14) Allyl chloride	2.666	41	113901	42.390	ug/l #	88
15) Acrylonitrile	3.068	53	251176	239.076	ug/l	99
16) Acetone	2.392	43	207742	240.142	ug/l	99
17) Carbon Disulfide	2.514	76	232591	44.267	ug/l	99
18) Methyl Acetate	2.709	43	104004	41.838	ug/l #	84
19) Methyl tert-butyl Ether	3.123	73	300104	48.165	ug/l #	89
20) Methylene Chloride	2.788	84	104656	49.706	ug/l #	84
21) trans-1,2-Dichloroethene	3.093	96	102795	45.660	ug/l	92
22) Diisopropyl ether	3.769	45	237088	43.791	ug/l #	79
23) Vinyl Acetate	3.727	43	925557	211.648	ug/l #	90
24) 1,1-Dichloroethane	3.611	63	164797	45.173	ug/l	98
25) 2-Butanone	4.568	43	317869	238.010	ug/l #	86
26) 2,2-Dichloropropane	4.477	77	107574	36.952	ug/l	99
27) cis-1,2-Dichloroethene	4.495	96	119582	47.392	ug/l	91
28) Bromochloromethane	4.910	49	45637	42.223	ug/l #	54
29) Tetrahydrofuran	5.013	42	192325	223.373	ug/l #	76
30) Chloroform	5.098	83	187845	46.954	ug/l	97
31) Cyclohexane	5.476	56	132083	43.149	ug/l	91
32) 1,1,1-Trichloroethane	5.391	97	168314	49.079	ug/l	95
36) 1,1-Dichloropropene	5.696	75	131998	46.422	ug/l	97
37) Ethyl Acetate	4.721	43	109116	44.167	ug/l #	93
38) Carbon Tetrachloride	5.684	117	150369	50.946	ug/l	97
39) Methylcyclohexane	7.385	83	160550	45.559	ug/l	91
40) Benzene	6.043	78	400762	49.091	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
 Data File : VX031078.D
 Acq On : 02 Sep 2022 19:32
 Operator : JC/MD
 Sample : N4297-18MSD
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-MW01D2-20220821MSD

Quant Time: Sep 03 00:51:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	62341	47.330	ug/l #	86
42) 1,2-Dichloroethane	6.092	62	136971	50.863	ug/l	98
43) Isopropyl Acetate	6.348	43	175455	45.441	ug/l #	89
44) Trichloroethene	7.129	130	136474	52.816	ug/l	99
45) 1,2-Dichloropropane	7.433	63	93801	44.827	ug/l	98
46) Dibromomethane	7.586	93	77028	49.183	ug/l #	87
47) Bromodichloromethane	7.824	83	146743	50.152	ug/l	99
48) Methyl methacrylate	7.696	41	91343	46.859	ug/l #	84
49) 1,4-Dioxane	7.689	88	64383	1085.890	ug/l #	88
51) 4-Methyl-2-Pentanone	8.580	43	592376	238.652	ug/l	88
52) Toluene	8.720	92	262568	49.595	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	146539	50.580	ug/l	98
54) cis-1,3-Dichloropropene	8.372	75	160376	49.506	ug/l #	88
55) 1,1,2-Trichloroethane	9.153	97	113638	50.489	ug/l	96
56) Ethyl methacrylate	9.122	69	156002	50.701	ug/l #	84
57) 1,3-Dichloropropane	9.311	76	172477	48.466	ug/l	99
59) 2-Hexanone	9.433	43	465883	248.197	ug/l	87
60) Dibromochloromethane	9.525	129	125769	53.815	ug/l	98
61) 1,2-Dibromoethane	9.610	107	123961	52.169	ug/l	100
64) Tetrachloroethene	9.275	164	130215	50.850	ug/l	96
65) Chlorobenzene	10.079	112	295587	48.257	ug/l	97
66) 1,1,1,2-Tetrachloroethane	10.165	131	113829	50.328	ug/l	98
67) Ethyl Benzene	10.195	91	512189	48.809	ug/l	99
68) m/p-Xylenes	10.305	106	403109	97.944	ug/l	99
69) o-Xylene	10.646	106	197434	48.913	ug/l	98
70) Styrene	10.658	104	335345	51.337	ug/l	100
71) Bromoform	10.799	173	97153	55.242	ug/l	100
73) Isopropylbenzene	10.963	105	526546	49.568	ug/l	100
74) N-amyl acetate	10.841	43	136258	42.495	ug/l #	82
75) 1,1,2,2-Tetrachloroethane	11.213	83	155768	45.855	ug/l	97
76) 1,2,3-Trichloropropane	11.244	75	179355	57.576	ug/l	88
77) Bromobenzene	11.201	156	135841	49.599	ug/l	87
78) n-propylbenzene	11.305	91	593949	49.092	ug/l	98
79) 2-Chlorotoluene	11.366	91	353597	47.463	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	444365	49.709	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	46811	47.227	ug/l	92
82) 4-Chlorotoluene	11.457	91	403188	48.368	ug/l	97
83) tert-Butylbenzene	11.719	119	429019	48.220	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	448247	50.450	ug/l	98
85) sec-Butylbenzene	11.890	105	545470	46.807	ug/l	98
86) p-Isopropyltoluene	12.012	119	459937	49.835	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	252000	48.035	ug/l	98
88) 1,4-Dichlorobenzene	12.042	146	252738	48.374	ug/l	98
89) n-Butylbenzene	12.335	91	382462	45.960	ug/l	97
90) Hexachloroethane	12.542	117	79719	47.663	ug/l	85
91) 1,2-Dichlorobenzene	12.335	146	250917	46.633	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	39950	48.269	ug/l	82
93) 1,2,4-Trichlorobenzene	13.591	180	190719	56.006	ug/l	97
94) Hexachlorobutadiene	13.725	225	74772	50.150	ug/l	99
95) Naphthalene	13.780	128	662549	55.470	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	178447	49.956	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
Data File : VX031078.D
Acq On : 02 Sep 2022 19:32
Operator : JC/MD
Sample : N4297-18MSD
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Quant Time: Sep 03 00:51:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#)=qualifier out of range (m)=manual integration (+)=signals summed						

Quant Time: Sep 03 00:51:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion:168 Resp: 174156
Ion Ratio Lower Upper
168 100
99 50.1 44.0 66.0



#2
Dichlorodifluoromethane
Concen: 42.189 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion: 85 Resp: 75602
Ion Ratio Lower Upper
85 100
87 32.4 16.5 49.5









#3
Chloromethane
Concen: 39.204 ug/l
RT: 1.288 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion: 50 Resp: 59142
Ion Ratio Lower Upper
50 100
52 33.1 26.3 39.5



#4
Vinyl Chloride
Concen: 41.852 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion: 62 Resp: 78343
Ion Ratio Lower Upper
62 100
64 32.0 25.2 37.8









#5
Bromomethane
Concen: 53.717 ug/l
RT: 1.611 min Scan# 81
Delta R.T. 0.006 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion: 94 Resp: 44954
Ion Ratio Lower Upper
94 100
96 100.7 76.9 115.3



#6
Chloroethane
Concen: 28.307 ug/l
RT: 1.684 min Scan# 98
Delta R.T. -0.001 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion: 64 Resp: 31966
Ion Ratio Lower Upper
64 100
66 34.5 27.2 40.8









#7
 Trichlorofluoromethane
 Concen: 48.590 ug/l
 RT: 1.886 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX031078.D
 Acq: 02 Sep 2022 19:32




Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-MW01D2-20220821MSD


Tgt Ion:101 Resp: 153695
 Ion Ratio Lower Upper
 101 100
 103 63.6 49.7 74.5



#8
 Diethyl Ether
 Concen: 45.123 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. -0.000 min
 Lab File: VX031078.D
 Acq: 02 Sep 2022 19:32

Tgt Ion: 74 Resp: 49778
 Ion Ratio Lower Upper
 74 100
 45 72.4 51.5 154.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 44.572 ug/l
RT: 2.331 min Scan# 204
Delta R.T. 0.006 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion:101 Resp: 93384
Ion Ratio Lower Upper
101 100
85 41.9 34.0 51.0
151 82.5 59.4 89.0



#10
Methyl Iodide
Concen: 54.546 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion:142 Resp: 111172
Ion Ratio Lower Upper
142 100
127 41.3 30.2 45.2
141 13.8 11.3 16.9









#11
Tert butyl alcohol
Concen: 275.346 ug/l
RT: 3.007 min Scan# 313
Delta R.T. 0.018 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion: 59 Resp: 123684
Ion Ratio Lower Upper
59 100
57 9.7 7.4 11.2



#12
1,1-Dichloroethene
Concen: 47.477 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion: 96 Resp: 96057
Ion Ratio Lower Upper
96 100
61 133.6 117.4 176.0
98 63.5 52.2 78.4









#13
Acrolein
Concen: 245.611 ug/l
RT: 2.239 min Scan# 189
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion: 56 Resp: 100715
Ion Ratio Lower Upper
56 100
55 71.6 56.6 84.8



#14
Allyl chloride
Concen: 42.390 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion: 41 Resp: 113901
Ion Ratio Lower Upper
41 100
39 66.4 48.9 73.3
76 45.7 26.9 40.3#









#15
Acrylonitrile
Concen: 239.076 ug/l
RT: 3.068 min Scan# 31
Delta R.T. -0.001 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion: 53 Resp: 251176
Ion Ratio Lower Upper
53 100
52 82.6 65.4 98.0
51 35.6 27.7 41.5



#16
Acetone
Concen: 240.142 ug/l
RT: 2.392 min Scan# 214
Delta R.T. 0.006 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion: 43 Resp: 207742
Ion Ratio Lower Upper
43 100
58 34.1 27.7 41.5









#17
Carbon Disulfide
Concen: 44.267 ug/l
RT: 2.514 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD

Tgt Ion: 76 Resp: 232591
Ion Ratio Lower Upper
76 100
78 9.3 7.8 11.6



#18
Methyl Acetate
Concen: 41.838 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion: 43 Resp: 104004
Ion Ratio Lower Upper
43 100
74 32.1 19.4 29.2#





#19

Methyl tert-butyl Ether

Concen: 48.165 ug/l

RT: 3.123 min Scan# 31

Delta R.T. 0.006 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32

Instrument :

MSVOA_X

ClientSampleId :





RW8-MW01D2-20220821MSD

Tgt Ion: 73 Resp: 300104

Ion Ratio Lower Upper

73 100

57 18.4 19.0 28.6#



#20

Methylene Chloride

Concen: 49.706 ug/l

RT: 2.788 min Scan# 279

Delta R.T. -0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32

Tgt Ion: 84 Resp: 104656




Ion Ratio Lower Upper


84 100

49 90.1 93.8 140.6#

51 28.1 29.4 44.0#

86 65.1 51.0 76.4









#21
trans-1,2-Dichloroethene
Concen: 45.660 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD

Tgt Ion: 96 Resp: 102795
Ion Ratio Lower Upper
96 100
61 120.0 105.6 158.4
98 64.6 50.1 75.1



#22
Diisopropyl ether
Concen: 43.791 ug/l
RT: 3.769 min Scan# 440
Delta R.T. -0.001 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion: 45 Resp: 237088
Ion Ratio Lower Upper
45 100
43 61.1 65.4 98.0#
87 37.9 22.3 33.5#
59 13.2 9.0 13.4









#23
 Vinyl Acetate
 Concen: 211.648 ug/l
 RT: 3.727 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX031078.D
 Acq: 02 Sep 2022 19:32




Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-MW01D2-20220821MSD


Tgt Ion: 43 Resp: 925557
 Ion Ratio Lower Upper
 43 100
 86 14.4 8.5 12.7#



#24
 1,1-Dichloroethane
 Concen: 45.173 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. -0.000 min
 Lab File: VX031078.D
 Acq: 02 Sep 2022 19:32

Tgt Ion: 63 Resp: 164797
 Ion Ratio Lower Upper
 63 100
 98 8.0 3.5 10.5
 100 5.1 2.3 6.9









#25
2-Butanone
Concen: 238.010 ug/l
RT: 4.568 min Scan# 51
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion: 43 Resp: 317869
Ion Ratio Lower Upper
43 100
72 34.5 21.7 32.5#



#26
2,2-Dichloropropane
Concen: 36.952 ug/l
RT: 4.477 min Scan# 556
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion: 77 Resp: 107574
Ion Ratio Lower Upper
77 100
97 26.5 13.0 39.0





#27

cis-1,2-Dichloroethene

Concen: 47.392 ug/l

RT: 4.495 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32

Instrument :

MSVOA_X

ClientSampleId :

RW8-MW01D2-20220821MSD





Tgt Ion: 96 Resp: 119582

Ion Ratio Lower Upper

96 100

61 119.8 0.0 270.0

98 65.1 0.0 127.8



#28

Bromochloromethane

Concen: 42.223 ug/l

RT: 4.910 min Scan# 627

Delta R.T. 0.006 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32




Tgt Ion: 49 Resp: 45637


Ion Ratio Lower Upper

49 100

129 3.4 0.0 4.4

130 119.9 62.8 94.2#





#29

Tetrahydrofuran

Concen: 223.373 ug/l

RT: 5.013 min Scan# 644

Delta R.T. 0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32

Instrument :

MSVOA_X

ClientSampleId :

RW8-MW01D2-20220821MSD





Tgt Ion: 42 Resp: 192325

Ion Ratio Lower Upper

42 100

72 61.0 36.5 54.7#

71 57.9 33.9 50.9#



#30

Chloroform

Concen: 46.954 ug/l

RT: 5.098 min Scan# 658

Delta R.T. -0.001 min

Lab File: VX031078.D




Acq: 02 Sep 2022 19:32


Tgt Ion: 83 Resp: 187845

Ion Ratio Lower Upper

83 100

85 66.8 51.5 77.3









#31
Cyclohexane
Concen: 43.149 ug/l
RT: 5.476 min Scan# 712
Delta R.T. -0.001 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion: 56 Resp: 132083
Ion Ratio Lower Upper
56 100
69 36.0 25.4 38.0
84 94.5 68.9 103.3



#32
1,1,1-Trichloroethane
Concen: 49.079 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion: 97 Resp: 168314
Ion Ratio Lower Upper
97 100
99 65.8 51.4 77.0
61 37.9 35.8 53.8





#33

1,2-Dichloroethane-d4

Concen: 48.987 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32

Instrument :

MSVOA_X

ClientSampleId :





RW8-MW01D2-20220821MSD

Tgt Ion: 65 Resp: 109772

Ion Ratio Lower Upper

65 100

67 52.8 0.0 109.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32




Tgt Ion: 114 Resp: 278172


Ion Ratio Lower Upper

114 100

63 17.0 0.0 38.4

88 15.1 0.0 31.8





#35

Dibromofluoromethane

Concen: 51.877 ug/l

RT: 5.391 min Scan# 706

Delta R.T. 0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32

Instrument :

MSVOA_X

ClientSampleId :

RW8-MW01D2-20220821MSD





Tgt Ion:113 Resp: 105310

Ion Ratio Lower Upper

113 100

111 102.2 83.2 124.8

192 21.2 14.8 22.2



#36

1,1-Dichloropropene

Concen: 46.422 ug/l

RT: 5.696 min Scan# 756

Delta R.T. -0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32




Tgt Ion: 75 Resp: 131998


Ion Ratio Lower Upper

75 100

110 41.2 19.3 57.9

77 31.8 24.4 36.6





#37

Ethyl Acetate

Concen: 44.167 ug/l

RT: 4.721 min Scan# 596

Delta R.T. -0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32

Instrument :

MSVOA_X

ClientSampleId :

RW8-MW01D2-20220821MSD





Tgt Ion: 43 Resp: 109116

Ion Ratio Lower Upper

43 100

61 17.2 11.7 17.5

70 14.2 9.3 13.9#



#38

Carbon Tetrachloride

Concen: 50.946 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32




Tgt Ion: 117 Resp: 150369

Ion Ratio Lower Upper

117 100

119 96.3 74.1 111.1

121 31.1 25.8 38.8





#39

Methylcyclohexane

Concen: 45.559 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX031078.D


Acq: 02 Sep 2022 19:32

Instrument :

MSVOA_X

ClientSampleId :

RW8-MW01D2-20220821MSD






Tgt Ion: 83 Resp: 160550

Ion Ratio Lower Upper

83 100

55 61.7 57.7 86.5

98 48.2 36.5 54.7



#40

Benzene

Concen: 49.091 ug/l

RT: 6.043 min Scan# 813

Delta R.T. -0.001 min

Lab File: VX031078.D




Acq: 02 Sep 2022 19:32


Tgt Ion: 78 Resp: 400762

Ion Ratio Lower Upper

78 100

77 23.1 18.0 27.0









#41
Methacrylonitrile
Concen: 47.330 ug/l
RT: 4.928 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion:	41	Resp:	62341
Ion	Ratio	Lower	Upper
41	100		
39	55.0	42.9	64.3
67	95.2	60.1	90.1#
52	35.8	22.6	34.0#



#42
1,2-Dichloroethane
Concen: 50.863 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion:	62	Resp:	136971
Ion	Ratio	Lower	Upper
62	100		
98	11.0	0.0	20.2









#43
Isopropyl Acetate
Concen: 45.441 ug/l
RT: 6.348 min Scan# 863
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument : MSVOA_X
ClientSampleId : RW8-MW01D2-20220821MSD


Tgt Ion: 43 Resp: 175455
Ion Ratio Lower Upper
43 100
61 24.4 16.5 24.7
87 19.8 11.1 16.7#



#44
Trichloroethene
Concen: 52.816 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion:130 Resp: 136474
Ion Ratio Lower Upper
130 100
95 96.3 0.0 194.8





#45

1,2-Dichloropropane

Concen: 44.827 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. -0.007 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32

Instrument :

MSVOA_X

ClientSampleId :





RW8-MW01D2-20220821MSD

Tgt Ion: 63 Resp: 93801

Ion Ratio Lower Upper

63 100

65 30.3 25.2 37.8



#46

Dibromomethane

Concen: 49.183 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32




Tgt Ion: 93 Resp: 77028


Ion Ratio Lower Upper

93 100

95 85.8 67.4 101.0

174 117.5 75.6 113.4#









#47
Bromodichloromethane
Concen: 50.152 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument : MSVOA_X
ClientSampleId : RW8-MW01D2-20220821MSD


Tgt Ion: 83 Resp: 146743
Ion Ratio Lower Upper
83 100
85 66.4 53.5 80.3
127 9.2 6.6 10.0



#48
Methyl methacrylate
Concen: 46.859 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion: 41 Resp: 91343
Ion Ratio Lower Upper
41 100
69 107.8 68.8 103.2#
39 58.6 44.6 66.8





#49

1,4-Dioxane

Concen: 1085.890 ug/l

RT: 7.689 min Scan# 1083

Delta R.T. 0.013 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32

Instrument :

MSVOA_X

ClientSampleId :

RW8-MW01D2-20220821MSD





Tgt Ion: 88 Resp: 64383

Ion Ratio Lower Upper

88 100

43 22.4 23.8 35.8#

58 57.8 53.8 80.6



#50

Toluene-d8

Concen: 49.882 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX031078.D




Acq: 02 Sep 2022 19:32


Tgt Ion: 98 Resp: 372675

Ion Ratio Lower Upper

98 100

100 65.9 52.3 78.5





#51

4-Methyl-2-Pentanone

Concen: 238.652 ug/l

RT: 8.580 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32

Instrument :

MSVOA_X

ClientSampleId :





RW8-MW01D2-20220821MSD

Tgt Ion: 43 Resp: 592376

Ion Ratio Lower Upper

43 100

58 46.6 31.4 47.2



#52

Toluene

Concen: 49.595 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. -0.000 min

Lab File: VX031078.D




Acq: 02 Sep 2022 19:32


Tgt Ion: 92 Resp: 262568

Ion Ratio Lower Upper

92 100

91 175.5 139.2 208.8









#53
 t-1,3-Dichloropropene
 Concen: 50.580 ug/l
 RT: 8.982 min Scan# 1295
 Delta R.T. -0.000 min
 Lab File: VX031078.D
 Acq: 02 Sep 2022 19:32




Instrument : MSVOA_X
 ClientSampleId : RW8-MW01D2-20220821MSD


Tgt Ion: 75 Resp: 146539
 Ion Ratio Lower Upper
 75 100
 77 32.0 26.3 39.5



#54
 cis-1,3-Dichloropropene
 Concen: 49.506 ug/l
 RT: 8.372 min Scan# 1195
 Delta R.T. 0.000 min
 Lab File: VX031078.D
 Acq: 02 Sep 2022 19:32

Tgt Ion: 75 Resp: 160376
 Ion Ratio Lower Upper
 75 100
 77 32.0 25.0 37.4
 39 34.7 37.5 56.3#





#55

1,1,2-Trichloroethane

Concen: 50.489 ug/l

RT: 9.153 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32

Instrument :

MSVOA_X

ClientSampleId :

RW8-MW01D2-20220821MSD

Tgt Ion: 97 Resp: 113638





Ion Ratio Lower Upper

97 100

83 84.2 68.9 103.3

85 54.3 44.0 66.0

99 67.0 48.3 72.5



#56

Ethyl methacrylate

Concen: 50.701 ug/l

RT: 9.122 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32




Tgt Ion: 69 Resp: 156002


Ion Ratio Lower Upper

69 100

41 51.1 52.3 78.5#

39 28.6 28.2 42.4









#57
1,3-Dichloropropane
Concen: 48.466 ug/l
RT: 9.311 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion: 76 Resp: 172477
Ion Ratio Lower Upper
76 100
78 33.2 25.9 38.9



#59
2-Hexanone
Concen: 248.197 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion: 43 Resp: 465883
Ion Ratio Lower Upper
43 100
58 64.5 27.6 82.8









#60
Dibromochloromethane
Concen: 53.815 ug/l
RT: 9.525 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion:129 Resp: 125769
Ion Ratio Lower Upper
129 100
127 79.0 38.5 115.5



#61
1,2-Dibromoethane
Concen: 52.169 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion:107 Resp: 123961
Ion Ratio Lower Upper
107 100
109 96.0 76.6 114.8





#62

4-Bromofluorobenzene

Concen: 52.542 ug/l

RT: 11.085 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32

Instrument :

MSVOA_X

ClientSampleId :

RW8-MW01D2-20220821MSD





Tgt Ion: 95 Resp: 138646

Ion Ratio Lower Upper

95 100

174 86.2 0.0 149.0

176 84.5 0.0 146.0



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. -0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32




Tgt Ion: 117 Resp: 256472


Ion Ratio Lower Upper

117 100

82 49.1 42.7 64.1

119 31.2 25.4 38.2









#64
Tetrachloroethene
Concen: 50.850 ug/l
RT: 9.275 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion:164 Resp: 130215
Ion Ratio Lower Upper
164 100
166 131.7 98.2 147.4
129 98.1 80.8 121.2
131 92.9 75.5 113.3



#65
Chlorobenzene
Concen: 48.257 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion:112 Resp: 295587
Ion Ratio Lower Upper
112 100
114 32.4 24.6 36.8









#66
1,1,1,2-Tetrachloroethane
Concen: 50.328 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion:131 Resp: 113829
Ion Ratio Lower Upper
131 100
133 96.3 46.9 140.8
119 66.8 34.4 103.1



#67
Ethyl Benzene
Concen: 48.809 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion: 91 Resp: 512189
Ion Ratio Lower Upper
91 100
106 32.2 25.5 38.3









#68
m/p-Xylenes
Concen: 97.944 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion:106 Resp: 403109
Ion Ratio Lower Upper
106 100
91 199.1 158.2 237.4



#69
o-Xylene
Concen: 48.913 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion:106 Resp: 197434
Ion Ratio Lower Upper
106 100
91 211.2 103.8 311.6





#70

Styrene

Concen: 51.337 ug/l

RT: 10.658 min Scan# 1570

Delta R.T. -0.001 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32

Instrument :

MSVOA_X

ClientSampleId :

RW8-MW01D2-20220821MSD





Tgt Ion:104 Resp: 335345

Ion Ratio Lower Upper

104 100

78 47.2 37.8 56.8

103 55.2 44.5 66.7



#71

Bromoform

Concen: 55.242 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.006 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32




Tgt Ion:173 Resp: 97153


Ion Ratio Lower Upper

173 100

175 47.7 23.8 71.2

254 0.1 0.1 0.1









#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion:152 Resp: 131776
Ion Ratio Lower Upper
152 100
115 85.1 40.5 121.5
150 178.0 0.0 347.0



#73
Isopropylbenzene
Concen: 49.568 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion:105 Resp: 526546
Ion Ratio Lower Upper
105 100
120 26.7 13.5 40.4









#74
N-amy1 acetate
Concen: 42.495 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.007 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument : MSVOA_X
ClientSampleId : RW8-MW01D2-20220821MSD


Tgt Ion: 43 Resp: 136258
Ion Ratio Lower Upper
43 100
70 55.1 32.3 48.5#
55 29.2 18.4 27.6#
61 30.6 19.2 28.8#



#75
1,1,2,2-Tetrachloroethane
Concen: 45.855 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion: 83 Resp: 155768
Ion Ratio Lower Upper
83 100
131 11.1 4.5 13.7
85 66.7 32.5 97.5









#76
1,2,3-Trichloropropane
Concen: 57.576 ug/l
RT: 11.244 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument : MSVOA_X
ClientSampleId : RW8-MW01D2-20220821MSD


Tgt Ion: 75 Resp: 179355
Ion Ratio Lower Upper
75 100
77 32.0 19.8 59.3



#77
Bromobenzene
Concen: 49.599 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion: 156 Resp: 135841
Ion Ratio Lower Upper
156 100
77 139.1 82.7 247.9
158 97.1 47.5 142.5









#78
n-propylbenzene
Concen: 49.092 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument : MSVOA_X
ClientSampleId : RW8-MW01D2-20220821MSD


Tgt Ion: 91 Resp: 593949
Ion Ratio Lower Upper
91 100
120 24.3 11.6 34.8



#79
2-Chlorotoluene
Concen: 47.463 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion: 91 Resp: 353597
Ion Ratio Lower Upper
91 100
126 36.7 17.4 52.3









#80
1,3,5-Trimethylbenzene
Concen: 49.709 ug/l
RT: 11.451 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion:105 Resp: 444365
Ion Ratio Lower Upper
105 100
120 48.5 24.4 73.2



#81
trans-1,4-Dichloro-2-butene
Concen: 47.227 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.006 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion: 75 Resp: 46811
Ion Ratio Lower Upper
75 100
53 85.3 74.2 111.4
89 49.8 34.8 52.2









#82
4-Chlorotoluene
Concen: 48.368 ug/l
RT: 11.457 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion: 91 Resp: 403188
Ion Ratio Lower Upper
91 100
126 32.3 15.3 45.9



#83
tert-Butylbenzene
Concen: 48.220 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion: 119 Resp: 429019
Ion Ratio Lower Upper
119 100
91 56.3 28.3 84.9
134 22.9 11.2 33.5









#84
1,2,4-Trimethylbenzene
Concen: 50.450 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion:105 Resp: 448247
Ion Ratio Lower Upper
105 100
120 44.9 22.9 68.8



#85
sec-Butylbenzene
Concen: 46.807 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.006 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion:105 Resp: 545470
Ion Ratio Lower Upper
105 100
134 21.0 10.0 30.0





#86

p-Isopropyltoluene

Concen: 49.835 ug/l

RT: 12.012 min Scan# 1792

Delta R.T. -0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32

Instrument :

MSVOA_X

ClientSampleId :

RW8-MW01D2-20220821MSD





Tgt Ion:119 Resp: 459937

Ion Ratio Lower Upper

119 100

134 26.9 13.9 41.6

91 22.9 11.8 35.3



#87

1,3-Dichlorobenzene

Concen: 48.035 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32




Tgt Ion:146 Resp: 252000


Ion Ratio Lower Upper

146 100

111 39.2 20.3 60.8

148 65.0 31.8 95.3









#88
1,4-Dichlorobenzene
Concen: 48.374 ug/l
RT: 12.042 min Scan# 11
Delta R.T. -0.001 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32




Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821MSD


Tgt Ion:146 Resp: 252738
Ion Ratio Lower Upper
146 100
111 37.0 20.3 60.9
148 64.7 32.3 96.8



#89
n-Butylbenzene
Concen: 45.960 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX031078.D
Acq: 02 Sep 2022 19:32

Tgt Ion: 91 Resp: 382462
Ion Ratio Lower Upper
91 100
92 53.4 26.2 78.5
134 27.9 12.6 37.8





#90

Hexachloroethane

Concen: 47.663 ug/l

RT: 12.542 min Scan# 1879

Delta R.T. 0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32

Instrument :

MSVOA_X

ClientSampleId :





RW8-MW01D2-20220821MSD

Tgt Ion:117 Resp: 79719

Ion Ratio Lower Upper

117 100

201 80.2 34.2 102.5



#91

1,2-Dichlorobenzene

Concen: 46.633 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32




Tgt Ion:146 Resp: 250917


Ion Ratio Lower Upper

146 100

111 40.0 20.7 62.1

148 64.1 32.0 96.2





#92

1,2-Dibromo-3-Chloropropane

Concen: 48.269 ug/l

RT: 12.945 min Scan# 1945

Delta R.T. -0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32

Instrument :

MSVOA_X

ClientSampleId :

RW8-MW01D2-20220821MSD





Tgt Ion: 75 Resp: 39950

Ion Ratio Lower Upper

75 100

155 105.5 45.0 134.8

157 134.2 56.9 170.7



#93

1,2,4-Trichlorobenzene

Concen: 56.006 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. -0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32




Tgt Ion: 180 Resp: 190719


Ion Ratio Lower Upper

180 100

182 93.1 48.2 144.6

145 30.9 15.9 47.6





#94

Hexachlorobutadiene

Concen: 50.150 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32

Instrument :

MSVOA_X

ClientSampleId :

RW8-MW01D2-20220821MSD





Tgt Ion:225 Resp: 74772

Ion Ratio Lower Upper

225 100

223 59.4 30.6 91.6

227 64.5 31.9 95.9



#95

Naphthalene

Concen: 55.470 ug/l

RT: 13.780 min Scan# 2082

Delta R.T. 0.006 min

Lab File: VX031078.D

Acq: 02 Sep 2022 19:32




Tgt Ion:128 Resp: 662549


Ion Ratio Lower Upper

128 100

127 12.6 10.2 15.2


129 10.9 8.6 13.0





#96
 1,2,3-Trichlorobenzene
 Concen: 49.956 ug/l
 RT: 13.963 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: VX031078.D
 Acq: 02 Sep 2022 19:32

Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-MW01D2-20220821MSD



Tgt Ion:180 Resp: 178447
 Ion Ratio Lower Upper
 180 100
 182 97.6 47.0 141.0
 145 35.4 17.0 50.9

