

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
 Data File : VX031058.D
 Acq On : 02 Sep 2022 11:44
 Operator : JC/MD
 Sample : N4298-02DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MDL01

Quant Time: Sep 03 00:48:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	232344	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	368371	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	332890	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	150872	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	140339	46.943	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.880%
35) Dibromofluoromethane	5.391	113	134823	50.153	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.300%
50) Toluene-d8	8.653	98	491567	49.685	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.360%
62) 4-Bromofluorobenzene	11.079	95	168884	48.329	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.660%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.495	96	11896	3.534	ug/l	84
44) Trichloroethene	7.129	130	104885	30.652	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
Data File : VX031058.D
Acq On : 02 Sep 2022 11:44
Operator : JC/MD
Sample : N4298-02DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MDL01

Quant Time: Sep 03 00:48:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

