

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
 Data File : VX031059.D
 Acq On : 02 Sep 2022 12:08
 Operator : JC/MD
 Sample : N4298-05DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MDL02

Quant Time: Sep 03 00:48:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	229015	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	361965	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	318680	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	143652	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	141594	48.051	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.100%
35) Dibromofluoromethane	5.391	113	135965	51.473	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.940%
50) Toluene-d8	8.653	98	491074	50.513	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.020%
62) 4-Bromofluorobenzene	11.079	95	166331	48.441	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.880%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	2151	0.781	ug/l	92
12) 1,1-Dichloroethene	2.319	96	932	0.350	ug/l #	65
27) cis-1,2-Dichloroethene	4.489	96	821	0.247	ug/l	47
44) Trichloroethene	7.129	130	158808	47.232	ug/l	98
64) Tetrachloroethene	9.281	164	1338	0.421	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
Data File : VX031059.D
Acq On : 02 Sep 2022 12:08
Operator : JC/MD
Sample : N4298-05DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MDL02

Quant Time: Sep 03 00:48:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

