

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
 Data File : VX031062.D
 Acq On : 02 Sep 2022 13:18
 Operator : JC/MD
 Sample : N4298-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MDL05

Quant Time: Sep 03 00:49:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	214544	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	341951	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	305411	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	137092	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	132351	47.944	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.880%
35) Dibromofluoromethane	5.391	113	126830	50.825	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.640%
50) Toluene-d8	8.653	98	457043	49.764	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.520%
62) 4-Bromofluorobenzene	11.079	95	153708	47.385	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.780%
Target Compounds						
					Qvalue	
3) Chloromethane	1.294	50	299	0.840	ug/l #	82
9) 1,1,2-Trichlorotrifluo...	2.331	101	3827	1.483	ug/l #	86
12) 1,1-Dichloroethene	2.325	96	547	0.219	ug/l	99
44) Trichloroethene	7.129	130	134089	42.215	ug/l	81
64) Tetrachloroethene	9.281	164	910	0.298	ug/l #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
Data File : VX031062.D
Acq On : 02 Sep 2022 13:18
Operator : JC/MD
Sample : N4298-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MDL05

Quant Time: Sep 03 00:49:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

