

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
 Data File : VX031065.D
 Acq On : 02 Sep 2022 14:28
 Operator : JC/MD
 Sample : N4298-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW181S-20220822

Quant Time: Sep 03 00:49:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	217353	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	352465	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	315451	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	142319	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	134590	48.125	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.260%
35) Dibromofluoromethane	5.391	113	133434	51.876	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.760%
50) Toluene-d8	8.653	98	473931	50.064	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.120%
62) 4-Bromofluorobenzene	11.085	95	159425	47.681	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.360%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.325	96	1110	0.440	ug/l #	75
24) 1,1-Dichloroethane	3.611	63	2387	0.524	ug/l #	90
27) cis-1,2-Dichloroethene	4.483	96	2234	0.709	ug/l	79
30) Chloroform	5.093	83	2357	0.472	ug/l #	81
32) 1,1,1-Trichloroethane	5.379	97	1387	0.324	ug/l #	35
44) Trichloroethene	7.123	130	3042	0.929	ug/l	91
64) Tetrachloroethene	9.275	164	5486	1.742	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
Data File : VX031065.D
Acq On : 02 Sep 2022 14:28
Operator : JC/MD
Sample : N4298-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW181S-20220822

Quant Time: Sep 03 00:49:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

