

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
 Data File : VX031067.D
 Acq On : 02 Sep 2022 15:15
 Operator : JC/MD
 Sample : N4298-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT162S1-20220822

Quant Time: Sep 03 00:50:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	214514	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	341007	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	302306	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135355	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	127496	46.192	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.380%
35) Dibromofluoromethane	5.391	113	126300	50.752	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.500%
50) Toluene-d8	8.653	98	448501	48.970	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.940%
62) 4-Bromofluorobenzene	11.079	95	152307	47.083	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.160%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	1217	0.488	ug/l #	88
24) 1,1-Dichloroethane	3.617	63	4176	0.929	ug/l	96
27) cis-1,2-Dichloroethene	4.507	96	723	0.233	ug/l	71
44) Trichloroethene	7.129	130	32423	10.236	ug/l	91
64) Tetrachloroethene	9.281	164	2434	0.806	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
Data File : VX031067.D
Acq On : 02 Sep 2022 15:15
Operator : JC/MD
Sample : N4298-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT162S1-20220822

Quant Time: Sep 03 00:50:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

