

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
 Data File : VX031068.D
 Acq On : 02 Sep 2022 15:38
 Operator : JC/MD
 Sample : N4297-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW202D1-20220820

Quant Time: Sep 03 00:50:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	212009	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	336698	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	303716	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138086	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	126867	46.507	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.020%
35) Dibromofluoromethane	5.385	113	125608	51.120	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.240%
50) Toluene-d8	8.653	98	445594	49.275	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.540%
62) 4-Bromofluorobenzene	11.085	95	150685	47.178	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.360%
Target Compounds						
					Qvalue	
19) Methyl tert-butyl Ether	3.117	73	6507	0.858	ug/l	93
24) 1,1-Dichloroethane	3.617	63	3000	0.676	ug/l #	90
27) cis-1,2-Dichloroethene	4.495	96	17396	5.663	ug/l	81
30) Chloroform	5.099	83	2354	0.483	ug/l #	75
44) Trichloroethene	7.129	130	14729	4.709	ug/l	99
64) Tetrachloroethene	9.275	164	14014	4.621	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
Data File : VX031068.D
Acq On : 02 Sep 2022 15:38
Operator : JC/MD
Sample : N4297-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW202D1-20220820

Quant Time: Sep 03 00:50:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

