

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
 Data File : VX031070.D
 Acq On : 02 Sep 2022 16:25
 Operator : JC/MD
 Sample : N4297-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-MW01D2-20220821

Quant Time: Sep 03 00:50:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	206637	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	326857	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	296346	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132710	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	125686	47.272	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.540%
35) Dibromofluoromethane	5.391	113	123208	51.653	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.300%
50) Toluene-d8	8.653	98	442648	50.423	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.840%
62) 4-Bromofluorobenzene	11.085	95	149771	48.304	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.600%
Target Compounds						
44) Trichloroethene	7.141	130	2542	0.837	ug/l	Qvalue 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
Data File : VX031070.D
Acq On : 02 Sep 2022 16:25
Operator : JC/MD
Sample : N4297-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220821

Quant Time: Sep 03 00:50:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

