

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
 Data File : VX031072.D
 Acq On : 02 Sep 2022 17:12
 Operator : JC/MD
 Sample : N4281-10 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP06-20220818

Quant Time: Sep 03 00:50:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	204265	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	327989	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	295157	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	142909	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	128248	48.796	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.600%
35) Dibromofluoromethane	5.391	113	123144	51.448	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.900%
50) Toluene-d8	8.653	98	446781	50.718	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.440%
62) 4-Bromofluorobenzene	11.085	95	156020	50.145	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.300%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.325	101	1738	0.707	ug/l #	79
27) cis-1,2-Dichloroethene	4.489	96	856	0.289	ug/l	67
44) Trichloroethene	7.129	130	98576	32.355	ug/l	87
64) Tetrachloroethene	9.281	164	2456	0.833	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
Data File : VX031072.D
Acq On : 02 Sep 2022 17:12
Operator : JC/MD
Sample : N4281-10 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP06-20220818

Quant Time: Sep 03 00:50:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

