

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
 Data File : VX031074.D
 Acq On : 02 Sep 2022 17:59
 Operator : JC/MD
 Sample : N4281-02DL 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D2-20220818DL

Quant Time: Sep 03 00:50:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	201903	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	328247	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	292838	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	147055	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	123993	47.729	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.460%
35) Dibromofluoromethane	5.391	113	121176	50.586	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.180%
50) Toluene-d8	8.653	98	433597	49.183	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.360%
62) 4-Bromofluorobenzene	11.085	95	150968	48.483	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.960%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.129	130	63476	20.818	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
Data File : VX031074.D
Acq On : 02 Sep 2022 17:59
Operator : JC/MD
Sample : N4281-02DL 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D2-20220818DL

Quant Time: Sep 03 00:50:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

