

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
 Data File : VX031079.D
 Acq On : 02 Sep 2022 19:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 03 00:51:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.514	0.453	11.9	79	0.00
3 P	Chloromethane	0.440	0.341	22.5	73	0.00
4 C	Vinyl Chloride	0.537	0.454	15.5#	74	0.00
5 T	Bromomethane	0.240	0.246	-2.5	89	0.00
6 T	Chloroethane	0.269	0.262	2.6	82	0.00
7 T	Trichlorofluoromethane	0.908	0.887	2.3	86	0.00
8 T	Diethyl Ether	0.317	0.276	12.9	77	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.602	0.560	7.0	84	0.00
10 T	Methyl Iodide	0.483	0.640	-32.5#	94	0.00
11 T	Tert butyl alcohol	0.153	0.125	18.3	81	0.00
12 CM	1,1-Dichloroethene	0.581	0.553	4.8#	81	0.00
13 T	Acrolein	0.118	0.110	6.8	80	0.00
14 T	Allyl chloride	0.771	0.656	14.9	74	0.00
15 T	Acrylonitrile	0.302	0.275	8.9	79	0.00
16 T	Acetone	0.248	0.222	10.5	81	0.00
17 T	Carbon Disulfide	1.509	1.329	11.9	79	0.00
18 T	Methyl Acetate	0.714	0.624	12.6	77	0.00
19 T	Methyl tert-butyl Ether	1.789	1.723	3.7	82	0.00
20 T	Methylene Chloride	0.693	0.599	13.6	81	0.00
21 T	trans-1,2-Dichloroethene	0.646	0.583	9.8	81	0.00
22 T	Diisopropyl ether	1.554	1.386	10.8	76	0.00
23 T	Vinyl Acetate	1.256	1.155	8.0	77	0.00
24 P	1,1-Dichloroethane	1.047	0.951	9.2	77	0.00
25 T	2-Butanone	0.383	0.341	11.0	77	0.00
26 T	2,2-Dichloropropane	0.836	0.685	18.1	73	0.00
27 T	cis-1,2-Dichloroethene	0.724	0.697	3.7	84	0.00
28 T	Bromochloromethane	0.310	0.262	15.5	76	0.00
29 T	Tetrahydrofuran	0.247	0.213	13.8	76	0.00
30 C	Chloroform	1.149	1.091	5.0#	84	0.00
31 T	Cyclohexane	0.879	0.768	12.6	74	0.00
32 T	1,1,1-Trichloroethane	0.985	0.984	0.1	85	0.00
33 S	1,2-Dichloroethane-d4	0.643	0.604	6.1	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	82	0.00
35 S	Dibromofluoromethane	0.365	0.378	-3.6	88	0.00
36 T	1,1-Dichloropropene	0.511	0.493	3.5	81	0.00
37 T	Ethyl Acetate	0.444	0.428	3.6	77	0.00
38 T	Carbon Tetrachloride	0.531	0.574	-8.1	87	0.00
39 T	Methylcyclohexane	0.633	0.609	3.8	76	0.00
40 TM	Benzene	1.467	1.481	-1.0	80	0.00
41 T	Methacrylonitrile	0.237	0.233	1.7	77	0.00
42 TM	1,2-Dichloroethane	0.484	0.502	-3.7	83	0.00
43 T	Isopropyl Acetate	0.694	0.668	3.7	77	0.00
44 TM	Trichloroethene	0.464	0.477	-2.8	81	0.00
45 C	1,2-Dichloropropane	0.376	0.359	4.5#	75	0.00
46 T	Dibromomethane	0.282	0.287	-1.8	84	0.00
47 T	Bromodichloromethane	0.526	0.540	-2.7	79	0.00
48 T	Methyl methacrylate	0.350	0.327	6.6	75	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
 Data File : VX031079.D
 Acq On : 02 Sep 2022 19:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 03 00:51:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.011	0.011	0.0	81	0.00
50 S	Toluene-d8	1.343	1.373	-2.2	83	0.00
51 T	4-Methyl-2-Pentanone	0.446	0.428	4.0	76	0.00
52 CM	Toluene	0.952	0.976	-2.5#	81	0.00
53 T	t-1,3-Dichloropropene	0.521	0.553	-6.1	79	0.00
54 T	cis-1,3-Dichloropropene	0.582	0.601	-3.3	78	0.00
55 T	1,1,2-Trichloroethane	0.405	0.424	-4.7	84	0.00
56 T	Ethyl methacrylate	0.553	0.584	-5.6	80	0.00
57 T	1,3-Dichloropropane	0.640	0.638	0.3	80	0.00
58 T	2-Chloroethyl Vinyl ether	0.276	0.264	4.3	70	-0.01
59 T	2-Hexanone	0.337	0.330	2.1	76	0.00
60 T	Dibromochloromethane	0.420	0.465	-10.7	82	0.00
61 T	1,2-Dibromoethane	0.427	0.468	-9.6	87	0.00
62 S	4-Bromofluorobenzene	0.474	0.498	-5.1	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.499	0.535	-7.2	88	0.00
65 PM	Chlorobenzene	1.194	1.180	1.2	82	0.00
66 T	1,1,1,2-Tetrachloroethane	0.441	0.461	-4.5	86	0.00
67 C	Ethyl Benzene	2.046	2.047	-0.0#	80	0.00
68 T	m/p-Xylenes	0.802	0.808	-0.7	82	0.00
69 T	o-Xylene	0.787	0.818	-3.9	84	0.00
70 T	Styrene	1.273	1.335	-4.9	82	0.00
71 P	Bromoform	0.343	0.391	-14.0	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
73 T	Isopropylbenzene	4.031	4.064	-0.8	79	0.00
74 T	N-amyl acetate	1.217	1.153	5.3	73	0.00
75 P	1,1,2,2-Tetrachloroethane	1.289	1.218	5.5	80	0.00
76 T	1,2,3-Trichloropropane	1.182	1.121	5.2	79	0.00
77 T	Bromobenzene	1.039	1.061	-2.1	84	0.00
78 T	n-propylbenzene	4.591	4.617	-0.6	80	0.00
79 T	2-Chlorotoluene	2.827	2.743	3.0	80	0.00
80 T	1,3,5-Trimethylbenzene	3.392	3.466	-2.2	83	0.00
81 T	trans-1,4-Dichloro-2-butene	0.376	0.367	2.4	74	0.00
82 T	4-Chlorotoluene	3.163	3.146	0.5	83	0.00
83 T	tert-Butylbenzene	3.376	3.353	0.7	79	0.00
84 T	1,2,4-Trimethylbenzene	3.371	3.457	-2.6	83	0.00
85 T	sec-Butylbenzene	4.422	4.292	2.9	75	0.00
86 T	p-Isopropyltoluene	3.502	3.631	-3.7	80	0.00
87 T	1,3-Dichlorobenzene	1.991	1.969	1.1	79	0.00
88 T	1,4-Dichlorobenzene	1.982	1.972	0.5	83	0.00
89 T	n-Butylbenzene	3.157	3.052	3.3	71	0.00
90 T	Hexachloroethane	0.635	0.635	0.0	78	0.00
91 T	1,2-Dichlorobenzene	2.042	1.959	4.1	75	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.314	0.306	2.5	78	0.00
93 T	1,2,4-Trichlorobenzene	1.292	1.504	-16.4	89	0.00
94 T	Hexachlorobutadiene	0.566	0.635	-12.2	93	0.00
95 T	Naphthalene	4.532	5.088	-12.3	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090222\
Data File : VX031079.D
Acq On : 02 Sep 2022 19:56
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 03 00:51:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.355	1.557	-14.9	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6