

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090420\
 Data File : VX018255.D
 Acq On : 04 Sep 2020 13:05
 Operator : JC/SP
 Sample : VX0904WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0904WBL01

Quant Time: Sep 05 06:47:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X090420W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Sep 04 12:13:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	4.98	128	81353	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.83	114	412345	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.10	117	387308	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.03	65	197295	30.74	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	102.47%
60) 4-Bromofluorobenzene	11.12	95	169997	25.32	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	84.40%
63) Toluene-d8	8.70	98	494544	28.97	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	96.57%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090420\
Data File : VX018255.D
Acq On : 04 Sep 2020 13:05
Operator : JC/SP
Sample : VX0904WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0904WBL01

Quant Time: Sep 05 06:47:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X090420W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Sep 04 12:13:41 2020
Response via : Initial Calibration

