

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX090420\
 Data File : VX018261.D
 Acq On : 04 Sep 2020 15:36
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Sep 05 07:01:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X090420W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Sep 04 12:13:41 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	97	0.00
2 M	Dichlorodifluoromethane	1.587	1.546	2.6	102	0.00
3 M	Chloromethane	1.686	1.649	2.2	103	0.00
4 M	Vinyl Chloride	1.776	1.729	2.6	102	0.00
5 M	Bromomethane	1.208	0.919	23.9	73	0.00
6 M	Chloroethane	1.079	1.072	0.6	102	0.00
7 M	Trichlorofluoromethane	3.009	2.924	2.8	103	0.00
8 T	Diethyl Ether	0.949	0.938	1.2	103	0.00
9	1,1,2-Trichlorotrifluoroeth	1.537	1.477	3.9	100	0.00
10 M	1,1-Dichloroethene	1.509	1.505	0.3	105	0.00
11	Methyl Iodide	1.623	0.836	48.5#	67	0.00
12	Methyl Acetate	1.987	2.072	-4.3	107	0.00
13 M	Acrolein	0.160	0.130	18.8	96	0.00
14 M	Acrylonitrile	0.900	0.925	-2.8	108	0.00
15 M	Acetone	0.245	0.260	-6.1	109	0.00
16 M	Carbon Disulfide	4.394	4.264	3.0	100	0.00
17	Allyl chloride	2.697	2.639	2.2	104	0.00
18 M	Methylene Chloride	1.771	1.678	5.3	97	0.00
19 M	trans-1,2-Dichloroethene	1.676	1.614	3.7	102	0.00
20 T	Diisopropyl ether	5.340	5.318	0.4	103	0.00
21 M	1,1-Dichloroethane	3.065	3.031	1.1	102	0.00
22 M	cis-1,2-Dichloroethene	1.854	1.837	0.9	103	0.00
23 M	tert-Butyl Alcohol	0.390	0.403	-3.3	113	0.00
24 M	Methyl tert-Butyl Ether	5.365	5.304	1.1	105	0.00
25 M	Chloroform	3.271	3.197	2.3	101	0.00
26	Cyclohexane	2.474	2.378	3.9	103	0.00
27 s	1,2-Dichloroethane-d4	2.367	2.330	1.6	94	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
29	1,1-Dichloropropene	0.458	0.435	5.0	101	0.00
30 M	2-Butanone	0.251	0.258	-2.8	107	0.00
31	2,2-Dichloropropane	0.553	0.508	8.1	97	0.00
32 M	1,1,1-Trichloroethane	0.599	0.600	-0.2	104	0.00
33 M	Carbon Tetrachloride	0.559	0.540	3.4	100	0.00
34 M	Benzene	1.307	1.284	1.8	102	0.00
35	Methacrylonitrile	0.264	0.276	-4.5	112	0.00
36 M	1,2-Dichloroethane	0.541	0.537	0.7	104	0.00
37 M	Trichloroethene	0.378	0.371	1.9	105	0.00
38	Methylcyclohexane	0.488	0.465	4.7	107	0.00
39 M	1,2-Dichloropropane	0.350	0.345	1.4	102	0.00
40	Dibromomethane	0.253	0.254	-0.4	104	0.00
41 M	Bromodichloromethane	0.524	0.510	2.7	103	0.00
42 M	Vinyl Acetate	0.935	0.941	-0.6	105	0.00
43	Ethyl Acetate	0.499	0.507	-1.6	106	0.00
44	Isopropyl Acetate	0.804	0.793	1.4	105	0.00
45 T	1,4-Dioxane	0.007	0.006#	14.3	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX090420\
 Data File : VX018261.D
 Acq On : 04 Sep 2020 15:36
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Sep 05 07:01:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X090420W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Sep 04 12:13:41 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.395	0.388	1.8	110	0.00
47	n-amyl Acetate	0.706	0.691	2.1	109	0.00
48 M	t-1,3-Dichloropropene	0.563	0.556	1.2	106	0.00
49 T	cis-1,3-Dichloropropene	0.591	0.576	2.5	102	0.00
50 M	1,1,2-Trichloroethane	0.352	0.354	-0.6	105	0.00
51	Ethyl methacrylate	0.510	0.485	4.9	105	0.00
52	1,3-Dichloropropane	0.585	0.569	2.7	102	0.00
53 M	Dibromochloromethane	0.438	0.418	4.6	104	0.00
54 M	1,2-Dibromoethane	0.386	0.387	-0.3	107	0.00
55 M	2-Chloroethyl vinyl ether	0.267	0.270	-1.1	108	0.00
56 M	Bromoform	0.333	0.317	4.8	107	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
58 M	4-Methyl-2-Pentanone	0.516	0.527	-2.1	107	0.00
59 M	2-Hexanone	0.391	0.396	-1.3	108	0.00
60 S	4-Bromofluorobenzene	0.520	0.510	1.9	96	0.00
61 M	Tetrachloroethene	0.374	0.368	1.6	103	0.00
62 M	Toluene	1.510	1.484	1.7	104	0.00
63 S	Toluene-d8	1.322	1.328	-0.5	97	0.00
64 M	Chlorobenzene	0.970	0.942	2.9	102	0.00
65	1,1,1,2-Tetrachloroethane	0.393	0.392	0.3	107	0.00
66 M	Ethyl Benzene	1.668	1.641	1.6	106	0.00
67 M	m/p-Xylenes	0.638	0.629	1.4	105	0.00
68 M	o-Xylene	0.603	0.582	3.5	107	0.00
69 M	Styrene	1.052	1.025	2.6	106	0.00
70	Isopropylbenzene	1.676	1.620	3.3	104	0.00
71 M	1,1,2,2-Tetrachloroethane	0.548	0.547	0.2	106	0.00
72	1,2,3-Trichloropropane	0.526	0.518	1.5	103	0.00
73	Bromobenzene	0.459	0.437	4.8	102	0.00
74	n-propylbenzene	1.974	1.871	5.2	102	0.00
75	2-Chlorotoluene	1.196	1.149	3.9	103	0.00
76	1,3,5-Trimethylbenzene	1.460	1.424	2.5	106	0.00
77	t-1,4-Dichloro-2-butene	0.205	0.180	12.2	98	0.00
78	4-Chlorotoluene	1.427	1.359	4.8	103	0.00
79	tert-butylbenzene	1.407	1.375	2.3	106	0.00
80	1,2,4-Trimethylbenzene	1.444	1.405	2.7	104	0.00
81	sec-Butylbenzene	1.643	1.521	7.4	102	0.00
82	p-Isopropyltoluene	1.532	1.436	6.3	103	0.00
83 M	1,3-Dichlorobenzene	0.821	0.798	2.8	103	0.00
84 M	1,4-Dichlorobenzene	0.821	0.793	3.4	102	0.00
85	n-Butylbenzene	1.349	1.246	7.6	104	0.00
86 T	Hexachloroethane	0.302	0.279	7.6	102	0.00
87 M	1,2-Dichlorobenzene	0.769	0.759	1.3	106	0.00
88	1,2-Dibromo-3-Chloropropane	0.140	0.139	0.7	113	0.00
89	1,2,4-Trichlorobenzene	0.520	0.485	6.7	105	0.00
90	Hexachlorobutadiene	0.220	0.195	11.4	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX090420\
Data File : VX018261.D
Acq On : 04 Sep 2020 15:36
Operator : JC/SP
Sample : VSTDCCC020
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC020

Quant Time: Sep 05 07:01:42 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X090420W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Sep 04 12:13:41 2020
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	1.624	1.556	4.2	107	0.00
92	1,2,3-Trichlorobenzene	0.515	0.499	3.1	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0