


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX090420\
 Data File : VX018248.D
 Acq On : 04 Sep 2020 09:50
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\624X090420W.M
 Title : SW846 624
 Last Update : Fri Aug 21 03:03:56 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.6	16667	PASS
75	95	30	60	51.2	45976	PASS
95	95	100	100	100.0	89786	PASS
96	95	5	9	6.0	5413	PASS
173	174	0.00	2	1.8	1315	PASS
174	95	50	100	81.0	72704	PASS
175	174	5	9	7.8	5640	PASS
176	174	95	101	95.8	69618	PASS
177	176	5	9	6.6	4592	PASS