

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091423\
 Data File : VX037721.D
 Acq On : 14 Sep 2023 11:39
 Operator : JC/MD
 Sample : VX0914MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0914MBL01

Quant Time: Sep 14 16:11:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	222789	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	367878	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	336714	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	147864	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	136291	37.811	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	75.620%#
35) Dibromofluoromethane	5.385	113	114367	43.860	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.720%
50) Toluene-d8	8.646	98	443083	45.055	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.120%#
62) 4-Bromofluorobenzene	11.079	95	153187	43.176	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091423\
Data File : VX037721.D
Acq On : 14 Sep 2023 11:39
Operator : JC/MD
Sample : VX0914MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0914MBL01

Quant Time: Sep 14 16:11:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

