

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091423\
 Data File : VX037726.D
 Acq On : 14 Sep 2023 13:42
 Operator : JC/MD
 Sample : 04374-06 20X
 Misc : 2.71g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DB-3-(31-33)ME

Quant Time: Sep 15 01:33:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	257036	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	426844	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	396953	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	184033	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	163930	39.419	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	78.840%
35) Dibromofluoromethane	5.385	113	134413	44.426	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.860%
50) Toluene-d8	8.647	98	513370	44.991	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.980%#
62) 4-Bromofluorobenzene	11.079	95	186224	45.237	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.480%
Target Compounds						
					Qvalue	
80) 1,3,5-Trimethylbenzene	11.451	105	20113	1.737	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	91067	7.800	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091423\
Data File : VX037726.D
Acq On : 14 Sep 2023 13:42
Operator : JC/MD
Sample : 04374-06 20X
Misc : 2.71g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DB-3-(31-33)ME

Quant Time: Sep 15 01:33:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

