

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091423\
 Data File : VX037730.D
 Acq On : 14 Sep 2023 15:50
 Operator : JC/MD
 Sample : 04408-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-17-0912

Quant Time: Sep 15 01:34:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	254739	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	436879	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	399161	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	168043	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	167525	40.647	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	81.300%
35) Dibromofluoromethane	5.385	113	140148	45.258	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.520%
50) Toluene-d8	8.647	98	523333	44.811	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.620%#
62) 4-Bromofluorobenzene	11.079	95	181436	43.062	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.120%
Target Compounds						
19) Methyl tert-butyl Ether	3.111	73	13268	1.282	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091423\
Data File : VX037730.D
Acq On : 14 Sep 2023 15:50
Operator : JC/MD
Sample : 04408-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-17-0912

Quant Time: Sep 15 01:34:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

