

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091423\
 Data File : VX037731.D
 Acq On : 14 Sep 2023 16:14
 Operator : JC/MD
 Sample : 04408-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-15-0912

Quant Time: Sep 15 01:34:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	257103	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	433289	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	395583	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	168567	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	167033	40.155	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	80.300%
35) Dibromofluoromethane	5.385	113	137485	44.766	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.540%
50) Toluene-d8	8.647	98	522544	45.114	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.220%#
62) 4-Bromofluorobenzene	11.079	95	179928	43.058	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.120%
Target Compounds						
19) Methyl tert-butyl Ether	3.111	73	2748	0.263	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091423\
Data File : VX037731.D
Acq On : 14 Sep 2023 16:14
Operator : JC/MD
Sample : 04408-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-15-0912

Quant Time: Sep 15 01:34:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

