

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091423\
 Data File : VX037732.D
 Acq On : 14 Sep 2023 16:39
 Operator : JC/MD
 Sample : 04390-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10A-GW-150-152

Quant Time: Sep 14 17:42:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	234925	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	405741	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	367213	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	150603	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	157923	41.549	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.100%
35) Dibromofluoromethane	5.385	113	128006	44.509	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.020%
50) Toluene-d8	8.647	98	482098	44.448	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.900%#
62) 4-Bromofluorobenzene	11.079	95	162264	41.467	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.940%#
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	5304	3.452	ug/l	89
19) Methyl tert-butyl Ether	3.117	73	17908	1.877	ug/l	94
22) Diisopropyl ether	3.763	45	6592	0.715	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091423\
Data File : VX037732.D
Acq On : 14 Sep 2023 16:39
Operator : JC/MD
Sample : 04390-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10A-GW-150-152

Quant Time: Sep 14 17:42:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

