

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091423\
 Data File : VX037735.D
 Acq On : 14 Sep 2023 17:52
 Operator : JC/MD
 Sample : 04390-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10A-TB-20230906

Quant Time: Sep 14 18:45:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 10:10:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	247748	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	409930	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	358033	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	143694	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	155073	38.687	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	77.380%#
35) Dibromofluoromethane	5.385	113	127806	43.986	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.980%
50) Toluene-d8	8.647	98	487491	44.486	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.980%#
62) 4-Bromofluorobenzene	11.079	95	157431	39.821	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	79.640%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091423\
Data File : VX037735.D
Acq On : 14 Sep 2023 17:52
Operator : JC/MD
Sample : 04390-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10A-TB-20230906

Quant Time: Sep 14 18:45:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 10:10:04 2023
Response via : Initial Calibration

