

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092022\
 Data File : VX031486.D
 Acq On : 20 Sep 2022 10:41
 Operator : JC/MD
 Sample : VX0920MBL01
 Misc : 5.0g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0920MBL01

Quant Time: Sep 20 14:47:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	51568	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	79104	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	85220	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54009	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	48967	51.718	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.440%
35) Dibromofluoromethane	5.385	113	40514	49.544	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.080%
50) Toluene-d8	8.647	98	147448	46.785	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.580%
62) 4-Bromofluorobenzene	11.079	95	46023	43.112	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092022\
Data File : VX031486.D
Acq On : 20 Sep 2022 10:41
Operator : JC/MD
Sample : VX0920MBL01
Misc : 5.0g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0920MBL01

Quant Time: Sep 20 14:47:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

