

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092022\
 Data File : VX031493.D
 Acq On : 20 Sep 2022 14:13
 Operator : JC/MD
 Sample : N4655-01
 Misc : 5.10G/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SU-3-091522

Quant Time: Sep 21 04:54:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	58130	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	89537	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	90087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55869	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	58445	54.825	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.640%	
35) Dibromofluoromethane	5.385	113	46312	50.043	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.080%	
50) Toluene-d8	8.646	98	176462	49.467	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.940%	
62) 4-Bromofluorobenzene	11.085	95	50911	42.133	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 84.260%	
Target Compounds						Qvalue
25) 2-Butanone	4.598	43	8115	9.860	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092022\
Data File : VX031493.D
Acq On : 20 Sep 2022 14:13
Operator : JC/MD
Sample : N4655-01
Misc : 5.10G/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SU-3-091522

Quant Time: Sep 21 04:54:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

