

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092022\
 Data File : VX031495.D
 Acq On : 20 Sep 2022 15:00
 Operator : JC/MD
 Sample : N4608-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOWER-1

Quant Time: Sep 20 17:19:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	49939	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	75680	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	79361	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51248	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	48396	52.805	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.600%
35) Dibromofluoromethane	5.391	113	39116	50.006	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.020%
50) Toluene-d8	8.653	98	147608	48.955	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.920%
62) 4-Bromofluorobenzene	11.079	95	45508	44.558	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.120%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	43670	113.082	ug/l	93
30) Chloroform	5.093	83	1453	0.854	ug/l	88
47) Bromodichloromethane	7.824	83	682	0.522	ug/l #	97
60) Dibromochloromethane	9.525	129	710	0.693	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092022\
Data File : VX031495.D
Acq On : 20 Sep 2022 15:00
Operator : JC/MD
Sample : N4608-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOWER-1

Quant Time: Sep 20 17:19:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

