

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092022\
 Data File : VX031496.D
 Acq On : 20 Sep 2022 15:23
 Operator : JC/MD
 Sample : N4608-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOWER-2

Quant Time: Sep 20 17:19:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	51861	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	81467	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	85722	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55205	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	51986	54.658	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.320%	
35) Dibromofluoromethane	5.385	113	43239	51.370	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.740%	
50) Toluene-d8	8.653	98	160883	49.568	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.140%	
62) 4-Bromofluorobenzene	11.079	95	48507	44.120	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.240%	
Target Compounds						
16) Acetone	2.386	43	42818	105.960	ug/l	100
60) Dibromochloromethane	9.525	129	1041	0.944	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092022\
Data File : VX031496.D
Acq On : 20 Sep 2022 15:23
Operator : JC/MD
Sample : N4608-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOWER-2

Quant Time: Sep 20 17:19:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

