

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092022\
 Data File : VX031503.D
 Acq On : 20 Sep 2022 18:07
 Operator : JC/MD
 Sample : N4668-06
 Misc : 5.70G/5mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 D22003-8

Quant Time: Sep 21 04:55:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	58972	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	91118	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	94951	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62507	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	57975	53.583	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.160%			
35) Dibromofluoromethane	5.385	113	45465	48.249	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.500%			
50) Toluene-d8	8.647	98	181665	50.042	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.080%			
62) 4-Bromofluorobenzene	11.085	95	54997	44.725	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 89.460%			
Target Compounds					Qvalue	
22) Diisopropyl ether	3.776	45	5874	1.812	ug/l #	78
31) Cyclohexane	5.477	56	12086	7.535	ug/l #	66
39) Methylcyclohexane	7.373	83	16478	9.247	ug/l	97
40) Benzene	6.037	78	5235	1.212	ug/l	98
67) Ethyl Benzene	10.195	91	16797	3.249	ug/l	96
68) m/p-Xylenes	10.305	106	3742	1.801	ug/l	88
69) o-Xylene	10.640	106	447	0.225	ug/l	65
73) Isopropylbenzene	10.963	105	5083	1.089	ug/l	97
78) n-propylbenzene	11.305	91	18999	3.538	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	19545	4.934	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	34977	8.966	ug/l	96
85) sec-Butylbenzene	11.890	105	6842	1.408	ug/l	98
86) p-Isopropyltoluene	12.012	119	2509	0.612	ug/l #	82
89) n-Butylbenzene	12.329	91	18350	5.250	ug/l #	42
95) Naphthalene	13.780	128	14621	3.263	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092022\
Data File : VX031503.D
Acq On : 20 Sep 2022 18:07
Operator : JC/MD
Sample : N4668-06
Misc : 5.70G/5mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
D22003-8

Quant Time: Sep 21 04:55:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

